

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083023\
 Data File : BN027317.D
 Acq On : 31 Aug 2023 20:04
 Operator : MA/JU
 Sample : PB155046BL
 Misc :
 ALS Vial : 98 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB155046BL

Quant Time: Sep 01 04:03:24 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 29 05:37:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.066	152	4621	0.400	ng	0.00
7) Naphthalene-d8	10.885	136	12191	0.400	ng	-0.01
13) Acenaphthene-d10	14.694	164	5354	0.400	ng	0.00
19) Phenanthrene-d10	17.443	188	14550	0.400	ng	0.00
29) Chrysene-d12	21.623	240	8573	0.400	ng	0.00
35) Perylene-d12	24.154	264	1912	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.582	112	8230	0.683	ng	0.00
5) Phenol-d6	7.199	99	8726	0.596	ng	0.00
8) Nitrobenzene-d5	9.251	82	4255	0.479	ng	-0.01
11) 2-Methylnaphthalene-d10	12.460	152	9082	0.507	ng	-0.01
14) 2,4,6-Tribromophenol	16.189	330	1371	0.692	ng	0.00
15) 2-Fluorobiphenyl	13.326	172	11229	0.552	ng	0.00
27) Fluoranthene-d10	19.463	212	19133	0.529	ng	0.00
31) Terphenyl-d14	20.048	244	12301	0.714	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083023\
Data File : BN027317.D
Acq On : 31 Aug 2023 20:04
Operator : MA/JU
Sample : PB155046BL
Misc :
ALS Vial : 98 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB155046BL

Quant Time: Sep 01 04:03:24 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 29 05:37:30 2023
Response via : Initial Calibration

