

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083023\
 Data File : BN027316.D
 Acq On : 31 Aug 2023 19:27
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 97 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 01 04:03:04 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 29 05:37:30 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	95	0.00
2	1,4-Dioxane	0.400	0.417	-4.2	93	0.00
3	n-Nitrosodimethylamine	0.400	0.449	-12.2	109	0.00
4 S	2-Fluorophenol	0.400	0.386	3.5	94	0.00
5 S	Phenol-d6	0.400	0.383	4.3	93	0.00
6	bis(2-Chloroethyl)ether	0.400	0.417	-4.2	98	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	106	0.00
8 S	Nitrobenzene-d5	0.400	0.398	0.5	101	0.00
9	Naphthalene	0.400	0.409	-2.2	109	0.01
10	Hexachlorobutadiene	0.400	0.406	-1.5	100	0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.398	0.5	102	0.00
12	2-Methylnaphthalene	0.400	0.402	-0.5	103	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	99	0.01
14 S	2,4,6-Tribromophenol	0.400	0.372	7.0	100	0.01
15 S	2-Fluorobiphenyl	0.400	0.405	-1.3	99	0.00
16	Acenaphthylene	0.400	0.395	1.3	100	0.00
17	Acenaphthene	0.400	0.434	-8.5	106	0.01
18	Fluorene	0.400	0.429	-7.2	104	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	100	0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.494	-23.5	136	0.00
21	4-Bromophenyl-phenylether	0.400	0.402	-0.5	104	0.00
22	Hexachlorobenzene	0.400	0.426	-6.5	108	0.01
23	Atrazine	0.400	0.373	6.8	99	0.00
24	Pentachlorophenol	0.400	0.352	12.0	98	0.00
25	Phenanthrene	0.400	0.411	-2.7	104	0.00
26	Anthracene	0.400	0.383	4.3	102	0.01
27 SURR	Fluoranthene-d10	0.400	0.388	3.0	97	0.00
28	Fluoranthene	0.400	0.390	2.5	99	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	86	0.00
30	Pyrene	0.400	0.452	-13.0	98	0.00
31 S	Terphenyl-d14	0.400	0.396	1.0	90	0.00
32	Benzo(a)anthracene	0.400	0.392	2.0	88	0.00
33	Chrysene	0.400	0.414	-3.5	86	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.357	10.8	76	0.00
35 I	Perylene-d12	0.400	0.400	0.0	80	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.351	12.3	79	0.01
37	Benzo(b)fluoranthene	0.400	0.405	-1.3	86	0.00
38	Benzo(k)fluoranthene	0.400	0.402	-0.5	86	0.00
39 C	Benzo(a)pyrene	0.400	0.392	2.0	83	0.00
40	Dibenzo(a,h)anthracene	0.400	0.351	12.3	79	0.02
41	Benzo(g,h,i)perylene	0.400	0.366	8.5	79	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4