

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083023\
 Data File : BN027332.D
 Acq On : 01 Sep 2023 05:49
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 113 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 01 06:19:34 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 29 05:37:30 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	0.00
2	1,4-Dioxane	0.488	0.485	0.6	107	0.00
3	n-Nitrosodimethylamine	0.518	0.567	-9.5	128	0.00
4 S	2-Fluorophenol	1.043	1.035	0.8	117	0.00
5 S	Phenol-d6	1.268	1.286	-1.4	119	0.00
6	bis(2-Chloroethyl)ether	1.226	1.258	-2.6	116	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	131	0.00
8 S	Nitrobenzene-d5	0.292	0.289	1.0	125	0.00
9	Naphthalene	1.091	1.114	-2.1	135	0.00
10	Hexachlorobutadiene	0.203	0.206	-1.5	123	0.00
11 SURR	2-Methylnaphthalene-d10	0.587	0.606	-3.2	131	0.00
12	2-Methylnaphthalene	0.776	0.804	-3.6	131	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	129	0.00
14 S	2,4,6-Tribromophenol	0.148	0.149	-0.7	140	0.00
15 S	2-Fluorobiphenyl	1.521	1.529	-0.5	128	0.00
16	Acenaphthylene	1.821	1.793	1.5	129	0.00
17	Acenaphthene	1.215	1.286	-5.8	134	0.00
18	Fluorene	1.620	1.698	-4.8	131	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	130	0.00
20	4,6-Dinitro-2-methylphenol	0.006	0.006	0.0	156#	0.00
21	4-Bromophenyl-phenylether	0.211	0.211	0.0	134	0.00
22	Hexachlorobenzene	0.250	0.261	-4.4	137	0.00
23	Atrazine	0.158	0.155	1.9	136	0.00
24	Pentachlorophenol	0.096	0.083	13.5	125	0.00
25	Phenanthrene	1.176	1.187	-0.9	133	0.00
26	Anthracene	0.999	0.969	3.0	134	0.00
27 SURR	Fluoranthene-d10	0.994	0.988	0.6	130	0.00
28	Fluoranthene	1.384	1.362	1.6	129	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	112	0.00
30	Pyrene	1.590	1.790	-12.6	126	0.00
31 S	Terphenyl-d14	0.804	0.841	-4.6	124	0.00
32	Benzo(a)anthracene	1.371	1.339	2.3	114	0.00
33	Chrysene	1.507	1.577	-4.6	113	0.00
34	Bis(2-ethylhexyl)phthalate	0.662	0.665	-0.5	111	0.00
35 I	Perylene-d12	1.000	1.000	0.0	98	0.00
36	Indeno(1,2,3-cd)pyrene	1.624	1.402	13.7	95	0.00
37	Benzo(b)fluoranthene	1.548	1.585	-2.4	106	0.00
38	Benzo(k)fluoranthene	1.591	1.585	0.4	105	0.00
39 C	Benzo(a)pyrene	1.447	1.378	4.8	99	0.00
40	Dibenzo(a,h)anthracene	1.273	1.106	13.1	96	0.00
41	Benzo(g,h,i)perylene	1.476	1.306	11.5	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4