

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083023\
 Data File : BN027358.D
 Acq On : 01 Sep 2023 23:29
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 139 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 02 01:56:07 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 01 09:08:53 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
2	1,4-Dioxane	0.488	0.493	-1.0	79	0.00
3	n-Nitrosodimethylamine	0.518	0.589	-13.7	96	0.00
4 S	2-Fluorophenol	1.043	0.974	6.6	79	0.00
5 S	Phenol-d6	1.268	1.211	4.5	81	0.00
6	bis(2-Chloroethyl)ether	1.226	1.237	-0.9	83	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
8 S	Nitrobenzene-d5	0.292	0.285	2.4	91	0.00
9	Naphthalene	1.091	1.089	0.2	97	0.01
10	Hexachlorobutadiene	0.203	0.202	0.5	89	0.00
11 SURR	2-Methylnaphthalene-d10	0.587	0.563	4.1	90	0.00
12	2-Methylnaphthalene	0.776	0.746	3.9	89	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	88	0.00
14 S	2,4,6-Tribromophenol	0.148	0.141	4.7	91	0.00
15 S	2-Fluorobiphenyl	1.521	1.463	3.8	84	0.00
16	Acenaphthylene	1.821	1.742	4.3	86	0.00
17	Acenaphthene	1.215	1.273	-4.8	91	0.00
18	Fluorene	1.620	1.649	-1.8	87	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	88	0.00
20	4,6-Dinitro-2-methylphenol	0.006	0.006	0.0	104	0.00
21	4-Bromophenyl-phenylether	0.211	0.201	4.7	87	0.00
22	Hexachlorobenzene	0.250	0.256	-2.4	91	0.00
23	Atrazine	0.158	0.142	10.1	84	0.00
24	Pentachlorophenol	0.096	0.084	12.5	85	0.00
25	Phenanthrene	1.176	1.163	1.1	88	0.00
26	Anthracene	0.999	0.947	5.2	88	0.00
27 SURR	Fluoranthene-d10	0.994	0.960	3.4	85	0.00
28	Fluoranthene	1.384	1.325	4.3	85	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	77	0.00
30	Pyrene	1.590	1.720	-8.2	84	0.00
31 S	Terphenyl-d14	0.804	0.804	0.0	82	0.00
32	Benzo(a)anthracene	1.371	1.313	4.2	77	0.00
33	Chrysene	1.507	1.526	-1.3	76	0.00
34	Bis(2-ethylhexyl)phthalate	0.662	0.720	-8.8	83	0.00
35 I	Perylene-d12	1.000	1.000	0.0	69	0.00
36	Indeno(1,2,3-cd)pyrene	1.624	1.468	9.6	70	0.00
37	Benzo(b)fluoranthene	1.548	1.632	-5.4	78	0.00
38	Benzo(k)fluoranthene	1.591	1.552	2.5	73	0.00
39 C	Benzo(a)pyrene	1.447	1.338	7.5	68	0.00
40	Dibenzo(a,h)anthracene	1.273	1.140	10.4	70	0.00
41	Benzo(g,h,i)perylene	1.476	1.353	8.3	69	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4