


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083023\
Data File : BN027267.D
Acq On : 30 Aug 2023 09:49
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Aug 29 05:37:30 2023






AutoFind: Scans 2536, 2537, 2538; Background Corrected with Scan 2529

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.6	162627	PASS
68	69	0.00	2	1.7	3096	PASS
69	198	0.00	100	43.2	186987	PASS
70	69	0.00	2	0.5	976	PASS
127	198	10	80	50.5	218475	PASS
197	198	0.00	2	0.5	2186	PASS
198	198	100	100	100.0	432789	PASS
199	198	5	9	6.8	29520	PASS
275	198	10	60	23.6	102259	PASS
365	198	1	100	3.0	12802	PASS
441	198	0.01	100	9.4	40605	PASS
442	442	50	100	100.0	256683	PASS
443	442	15	24	18.6	47680	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083023\
 Data File : BN027267.D
 Acq On : 30 Aug 2023 09:49
 Operator : MA/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Quant Time: Aug 31 06:24:33 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 31 06:24:29 2023
 Response via : Initial Calibration



TIC: BN027267.D\data.ms

(70) Pentachlorophenol (C)

17.069min (0.000) 21414.60 ng




response 476776

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	65.13
264.00	61.60	64.28
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083023\
Data File : BN027267.D
Acq On : 30 Aug 2023 09:49
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Aug 31 06:24:33 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 31 06:24:29 2023
Response via : Initial Calibration



TIC: BN027267.D\data.ms

(77) Benzidine

19.698min (0.000) 0.00 ng

response 3267448

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.48
183.00	13.20	11.74
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/30/2023	BNA_N	<u>BN027267.D</u>
Compound Name	Response	Retention Time
DDT	1573556	20.892
DDD	25579	20.451
DDE	1304	19.939
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
26883	1600439	1.68

Instrument :
BNA_N
ClientSampleId :
DFTPP