



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083023\
Data File : BN027333.D
Acq On : 01 Sep 2023 07:01
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 114 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri Sep 01 09:08:53 2023






AutoFind: Scans 2535, 2536, 2537; Background Corrected with Scan 2527

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
51	198	10	80	31.8	235614	PASS
68	69	0.00	2	1.7	4821	PASS
69	198	0.00	100	37.4	277134	PASS
70	69	0.00	2	0.6	1614	PASS
127	198	10	80	45.4	337003	PASS
197	198	0.00	2	0.3	2466	PASS
198	198	100	100	100.0	741675	PASS
199	198	5	9	6.8	50259	PASS
275	198	10	60	24.8	183616	PASS
365	198	1	100	3.0	22499	PASS
441	198	0.01	100	10.9	80504	PASS
442	442	50	100	100.0	464768	PASS
443	442	15	24	21.1	97885	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083023\
Data File : BN027333.D
Acq On : 01 Sep 2023 07:01
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 114 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Sep 01 07:35:50 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 01 07:34:32 2023
Response via : Initial Calibration



TIC: BN027333.D\data.ms

(70) Pentachlorophenol (C)

17.063min (-0.000) 32188.41 ng




response 1161485

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	62.35
264.00	61.60	63.98
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083023\
 Data File : BN027333.D
 Acq On : 01 Sep 2023 07:01
 Operator : MA/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 114 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Quant Time: Sep 01 07:35:50 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 01 07:34:32 2023
 Response via : Initial Calibration



TIC: BN027333.D\data.ms

(77) Benzidine

19.692min (-0.006) 328028.85 ng

response 6815226

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	15.17
183.00	13.20	12.10
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/30/2023	BNA_N	<u>BN027333.D</u>
Compound Name	Response	Retention Time
DDT	3194911	20.886
DDD	119991	20.445
DDE	2416	19.933
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
122407	3317318	3.69

Instrument :
BNA_N
ClientSampleId :
DFTPP