

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090119\
 Data File : BN007528.D
 Acq On : 01 Sep 2019 16:49
 Operator : HP/JU
 Sample : PB122534BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB122534BL

Quant Time: Sep 02 08:23:24 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 29 12:18:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.39	152	8520	0.40	ng	-0.02
7) Naphthalene-d8	10.15	136	34154	0.40	ng	-0.01
13) Acenaphthene-d10	14.04	164	21106	0.40	ng	-0.01
19) Phenanthrene-d10	16.79	188	52504	0.40	ng	-0.01
27) Chrysene-d12	21.01	240	49178	0.40	ng	-0.01
34) Perylene-d12	23.11	264	51500	0.40	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.03	112	8737	0.29	ng	0.00
5) Phenol-d6	6.59	99	10847	0.28	ng	0.00
8) Nitrobenzene-d5	8.53	82	7784	0.26	ng	0.00
11) 2-Methylnaphthalene-d10	11.75	152	18171	0.32	ng	0.00
14) 2,4,6-Tribromophenol	15.55	330	1849	0.17	ng	-0.01
15) 2-Fluorobiphenyl	12.66	172	23763	0.28	ng	0.00
25) Fluoranthene-d10	18.84	212	52682	0.31	ng	0.00
29) Terphenyl-d14	19.45	244	40703	0.32	ng	-0.01

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090119\
Data File : BN007528.D
Acq On : 01 Sep 2019 16:49
Operator : HP/JU
Sample : PB122534BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB122534BL

Quant Time: Sep 02 08:23:24 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 29 12:18:45 2019
Response via : Initial Calibration

