


Data Path : Z:\SVOASRV\HPCHEM1\BNA\_N\DATA\BN090119\  
 Data File : BN007526.D  
 Acq On : 01 Sep 2019 14:57  
 Operator : HP/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN082319.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Thu Aug 29 12:18:45 2019



AutoFind: Scans 2454, 2455, 2456; Background Corrected with Scan 2444

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 21.8      | 192827  | PASS             |
| 68          | 69           | 0.00         | 2            | 2.0       | 6612    | PASS             |
| 69          | 198          | 0.00         | 100          | 37.5      | 331606  | PASS             |
| 70          | 69           | 0.00         | 2            | 0.6       | 1953    | PASS             |
| 127         | 198          | 10           | 80           | 47.4      | 418664  | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 883568  | PASS             |
| 199         | 198          | 5            | 9            | 6.8       | 59741   | PASS             |
| 275         | 198          | 10           | 60           | 27.7      | 244650  | PASS             |
| 365         | 198          | 1            | 100          | 3.3       | 29162   | PASS             |
| 441         | 198          | 0.01         | 100          | 13.6      | 120560  | PASS             |
| 442         | 198          | 50           | 100          | 87.4      | 772522  | PASS             |
| 443         | 442          | 15           | 24           | 17.8      | 137501  | PASS             |