

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090221\
 Data File : BN016156.D
 Acq On : 03 Sep 2021 03:12
 Operator : CG/JU
 Sample : M3474-18
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AD1

Quant Time: Sep 03 06:16:08 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 02 13:32:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.867	152	2427	0.400	ng/ul	0.00
4) Naphthalene-d8	10.666	136	10048	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.506	164	6177	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.254	188	12571	0.400	ng/ul	0.00
17) Chrysene-d12	21.439	240	12408	0.400	ng/ul	0.00
23) Perylene-d12	23.833	264	12739	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.306	96	13706	4.852	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.255	152	4434	0.313	ng/ul	0.00
18) Fluoranthene-d10	19.285	212	13596	0.408	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090221\
Data File : BN016156.D
Acq On : 03 Sep 2021 03:12
Operator : CG/JU
Sample : M3474-18
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AD1

Quant Time: Sep 03 06:16:08 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 02 13:32:21 2021
Response via : Initial Calibration

