

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090222\
 Data File : BN021563.D
 Acq On : 02 Sep 2022 16:08
 Operator : CG/JU
 Sample : N4373-04
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE114D2-20220825

Quant Time: Sep 03 03:07:35 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 00:53:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.069	152	4568	0.400	ng	0.00
7) Naphthalene-d8	10.898	136	14424	0.400	ng	0.00
13) Acenaphthene-d10	14.718	164	7960	0.400	ng	0.00
19) Phenanthrene-d10	17.470	188	16358	0.400	ng	0.00
29) Chrysene-d12	21.664	240	11274	0.400	ng	0.00
35) Perylene-d12	24.182	264	8177	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.591	112	1232	0.138	ng	0.00
5) Phenol-d6	7.224	99	900	0.079	ng	0.00
8) Nitrobenzene-d5	9.243	82	2865	0.294	ng	0.00
11) 2-Methylnaphthalene-d10	12.486	152	10749	0.331	ng	0.00
14) 2,4,6-Tribromophenol	16.218	330	658	0.252	ng	0.00
15) 2-Fluorobiphenyl	13.349	172	12088	0.347	ng	0.00
27) Fluoranthene-d10	19.501	212	15889	0.378	ng	0.00
31) Terphenyl-d14	20.088	244	11772	0.465	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.403	88	18434	3.227	ng	Qvalue # 90
34) Bis(2-ethylhexyl)phtha...	21.556	149	1907	0.121	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090222\
Data File : BN021563.D
Acq On : 02 Sep 2022 16:08
Operator : CG/JU
Sample : N4373-04
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE114D2-20220825

Quant Time: Sep 03 03:07:35 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 02 00:53:31 2022
Response via : Initial Calibration

