

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090222\
 Data File : BN021568.D
 Acq On : 02 Sep 2022 19:15
 Operator : CG/JU
 Sample : N4454-05
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 OW-3B-49-54-083022-FD

Quant Time: Sep 03 03:09:08 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 00:53:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.069	152	4744	0.400	ng	0.00
7) Naphthalene-d8	10.898	136	15360	0.400	ng	0.00
13) Acenaphthene-d10	14.718	164	9301	0.400	ng	0.00
19) Phenanthrene-d10	17.470	188	20054	0.400	ng	0.00
29) Chrysene-d12	21.655	240	12770	0.400	ng	# 0.00
35) Perylene-d12	24.176	264	9514	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.592	112	1542	0.166	ng	0.00
5) Phenol-d6	7.217	99	1172	0.099	ng	0.00
8) Nitrobenzene-d5	9.243	82	3739	0.360	ng	0.00
11) 2-Methylnaphthalene-d10	12.486	152	12695	0.367	ng	0.00
14) 2,4,6-Tribromophenol	16.208	330	841	0.275	ng	-0.01
15) 2-Fluorobiphenyl	13.349	172	17392	0.428	ng	0.00
27) Fluoranthene-d10	19.497	212	17023	0.330	ng	0.00
31) Terphenyl-d14	20.088	244	13069	0.455	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.548	149	1702	0.095	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090222\
Data File : BN021568.D
Acq On : 02 Sep 2022 19:15
Operator : CG/JU
Sample : N4454-05
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
OW-3B-49-54-083022-FD

Quant Time: Sep 03 03:09:08 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 02 00:53:31 2022
Response via : Initial Calibration

