

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090222\
 Data File : BN021589.D
 Acq On : 03 Sep 2022 09:57
 Operator : CG/JU
 Sample : N4471-17
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EB01-083122

Quant Time: Sep 05 02:37:49 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 00:53:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.069	152	4990	0.400	ng	0.00
7) Naphthalene-d8	10.898	136	17264	0.400	ng	0.00
13) Acenaphthene-d10	14.718	164	14467	0.400	ng	0.00
19) Phenanthrene-d10	17.470	188	32690	0.400	ng	0.00
29) Chrysene-d12	21.655	240	21373	0.400	ng	# 0.00
35) Perylene-d12	24.173	264	15756	0.400	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.592	112	538	0.055	ng	0.00
5) Phenol-d6	7.217	99	1102	0.089	ng	0.00
8) Nitrobenzene-d5	9.233	82	3376	0.290	ng	-0.01
11) 2-Methylnaphthalene-d10	12.482	152	10534	0.271	ng	0.00
14) 2,4,6-Tribromophenol	16.218	330	47	0.010	ng	0.00
15) 2-Fluorobiphenyl	13.349	172	13025	0.206	ng	0.00
27) Fluoranthene-d10	19.497	212	16682	0.198	ng	0.00
31) Terphenyl-d14	20.088	244	15471	0.322	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.403	88	610	0.098	ng	# 66
34) Bis(2-ethylhexyl)phtha...	21.557	149	1382	0.046	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090222\
Data File : BN021589.D
Acq On : 03 Sep 2022 09:57
Operator : CG/JU
Sample : N4471-17
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EB01-083122

Quant Time: Sep 05 02:37:49 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 02 00:53:31 2022
Response via : Initial Calibration

