

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090222\
 Data File : BN021591.D
 Acq On : 03 Sep 2022 11:10
 Operator : CG/JU
 Sample : N4454-08
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW-13B-45-50-083022

Quant Time: Sep 05 02:38:25 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 00:53:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.069	152	4650	0.400	ng	0.00
7) Naphthalene-d8	10.898	136	15447	0.400	ng	0.00
13) Acenaphthene-d10	14.718	164	9394	0.400	ng	0.00
19) Phenanthrene-d10	17.470	188	20216	0.400	ng	0.00
29) Chrysene-d12	21.655	240	14087	0.400	ng	# 0.00
35) Perylene-d12	24.170	264	10360	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.592	112	1507	0.166	ng	0.00
5) Phenol-d6	7.217	99	1167	0.101	ng	0.00
8) Nitrobenzene-d5	9.243	82	3294	0.316	ng	0.00
11) 2-Methylnaphthalene-d10	12.482	152	10615	0.305	ng	0.00
14) 2,4,6-Tribromophenol	16.208	330	986	0.320	ng	-0.01
15) 2-Fluorobiphenyl	13.349	172	15906	0.387	ng	0.00
27) Fluoranthene-d10	19.497	212	18416	0.354	ng	0.00
31) Terphenyl-d14	20.088	244	14664	0.463	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.548	149	2239	0.113	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090222\
Data File : BN021591.D
Acq On : 03 Sep 2022 11:10
Operator : CG/JU
Sample : N4454-08
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW-13B-45-50-083022

Quant Time: Sep 05 02:38:25 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 02 00:53:31 2022
Response via : Initial Calibration

