

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090222\
 Data File : BN021573.D
 Acq On : 02 Sep 2022 22:20
 Operator : CG/JU
 Sample : PB147294BS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB147294BS

Quant Time: Sep 03 03:10:54 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 00:53:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.069	152	4546	0.400	ng	0.00
7) Naphthalene-d8	10.898	136	14681	0.400	ng	0.00
13) Acenaphthene-d10	14.718	164	8353	0.400	ng	0.00
19) Phenanthrene-d10	17.470	188	17321	0.400	ng	0.00
29) Chrysene-d12	21.655	240	10971	0.400	ng	# 0.00
35) Perylene-d12	24.173	264	7463	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.591	112	3877	0.436	ng	0.00
5) Phenol-d6	7.216	99	4848	0.428	ng	0.00
8) Nitrobenzene-d5	9.243	82	3913	0.395	ng	0.00
11) 2-Methylnaphthalene-d10	12.482	152	14401	0.435	ng	0.00
14) 2,4,6-Tribromophenol	16.208	330	1020	0.372	ng	-0.01
15) 2-Fluorobiphenyl	13.349	172	14481	0.397	ng	0.00
27) Fluoranthene-d10	19.497	212	16623	0.373	ng	0.00
31) Terphenyl-d14	20.088	244	10971	0.445	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.403	88	2412	0.424	ng	# 88
3) n-Nitrosodimethylamine	3.743	42	2132	0.416	ng	# 93
6) bis(2-Chloroethyl)ether	7.484	93	5867	0.411	ng	97
9) Naphthalene	10.951	128	17477	0.402	ng	99
10) Hexachlorobutadiene	11.229	225	3002	0.399	ng	# 100
12) 2-Methylnaphthalene	12.180	142	2404	0.459	ng	97
16) Acenaphthylene	14.440	152	14958	0.381	ng	100
17) Acenaphthene	14.782	154	10948	0.397	ng	99
18) Fluorene	15.765	166	13574	0.399	ng	99
21) 4-Bromophenyl-phenylether	16.659	248	4180	0.386	ng	99
22) Hexachlorobenzene	16.772	284	5186	0.391	ng	99
23) Atrazine	16.916	200	2317	0.366	ng	96
24) Pentachlorophenol	17.121	266	807	0.405	ng	96
25) Phenanthrene	17.511	178	23241	0.389	ng	100
26) Anthracene	17.603	178	18090	0.381	ng	99
28) Fluoranthene	19.527	202	22919	0.369	ng	99
30) Pyrene	19.893	202	24619	0.421	ng	96
32) Benzo(a)anthracene	21.637	228	15088	0.395	ng	99
33) Chrysene	21.691	228	18491	0.406	ng	99
34) Bis(2-ethylhexyl)phtha...	21.548	149	6756	0.439	ng	99
36) Indeno(1,2,3-cd)pyrene	26.822	276	13344	0.396	ng	100
37) Benzo(b)fluoranthene	23.398	252	14344	0.390	ng	99
38) Benzo(k)fluoranthene	23.448	252	14175	0.383	ng	99
39) Benzo(a)pyrene	24.059	252	10400	0.405	ng	98
40) Dibenzo(a,h)anthracene	26.845	278	10389	0.383	ng	96
41) Benzo(g,h,i)perylene	27.638	276	11215	0.377	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090222\
Data File : BN021573.D
Acq On : 02 Sep 2022 22:20
Operator : CG/JU
Sample : PB147294BS
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB147294BS


Quant Time: Sep 03 03:10:54 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 02 00:53:31 2022
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 8.069 min Scan# 60
 Delta R.T. -0.000 min
 Lab File: BN021573.D
 Acq: 02 Sep 2022 22:20


Instrument :
 BNA_N
 ClientSampleId :
 PB147294BS





Tgt Ion:152 Resp: 4546
 Ion Ratio Lower Upper
 152 100
 150 156.2 121.4 182.0
 115 60.5 48.6 72.8



#2
 1,4-Dioxane
 Concen: 0.424 ng
 RT: 3.403 min Scan# 17
 Delta R.T. -0.000 min
 Lab File: BN021573.D
 Acq: 02 Sep 2022 22:20



Tgt Ion: 88 Resp: 2412
 Ion Ratio Lower Upper
 88 100
 43 27.8 35.8 53.8#
 58 74.1 57.1 85.7








#3
 n-Nitrosodimethylamine
 Concen: 0.416 ng
 RT: 3.743 min Scan# 64
 Delta R.T. -0.000 min
 Lab File: BN021573.D
 Acq: 02 Sep 2022 22:20




Instrument :
 BNA_N
 ClientSampleId :
 PB147294BS

Tgt Ion: 42 Resp: 2132
 Ion Ratio Lower Upper
 42 100
 74 133.4 113.0 169.6
 44 17.6 10.5 15.7#



#4
 2-Fluorophenol
 Concen: 0.436 ng
 RT: 5.591 min Scan# 320
 Delta R.T. -0.000 min
 Lab File: BN021573.D
 Acq: 02 Sep 2022 22:20

Tgt Ion: 112 Resp: 3877
 Ion Ratio Lower Upper
 112 100
 64 56.8 43.5 65.3
 63 29.6 22.6 34.0








#5
Phenol-d6
Concen: 0.428 ng
RT: 7.216 min Scan# 54
Delta R.T. -0.000 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20




Instrument :
BNA_N
ClientSampleId :
PB147294BS

Tgt Ion: 99 Resp: 4848
Ion Ratio Lower Upper
99 100
42 19.4 15.5 23.3
71 31.6 24.2 36.2



#6
bis(2-Chloroethyl)ether
Concen: 0.411 ng
RT: 7.484 min Scan# 582
Delta R.T. -0.000 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20

Tgt Ion: 93 Resp: 5867
Ion Ratio Lower Upper
93 100
63 76.4 57.8 86.8
95 32.6 25.9 38.9








#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.898 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20




Instrument :
BNA_N
ClientSampleId :
PB147294BS


Tgt Ion:	136	Resp:	14681
Ion Ratio	Lower	Upper	
136	100		
137	11.3	9.3	13.9
54	7.6	6.9	10.3
68	6.7	6.1	9.1



#8
Nitrobenzene-d5
Concen: 0.395 ng
RT: 9.243 min Scan# 784
Delta R.T. -0.000 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20

Tgt Ion:	82	Resp:	3913
Ion Ratio	Lower	Upper	
82	100		
128	44.3	30.9	46.3
54	47.9	42.7	64.1









#9
Naphthalene
Concen: 0.402 ng
RT: 10.951 min Scan# 94
Delta R.T. -0.000 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20




Instrument :
BNA_N
ClientSampleId :
PB147294BS

Tgt Ion:128 Resp: 17477
Ion Ratio Lower Upper
128 100
129 11.3 9.3 13.9
127 13.0 11.0 16.4



#10
Hexachlorobutadiene
Concen: 0.399 ng
RT: 11.229 min Scan# 970
Delta R.T. -0.011 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20

Tgt Ion:225 Resp: 3002
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.0 51.2 76.8





#11

2-Methylnaphthalene-d10

Concen: 0.435 ng

RT: 12.482 min Scan# 1174

Delta R.T. -0.004 min

Lab File: BN021573.D

Acq: 02 Sep 2022 22:20

Instrument :

BNA_N

ClientSampleId :





PB147294BS

Tgt Ion:152 Resp: 14401

Ion Ratio Lower Upper

152 100

151 20.2 14.5 21.7



#12

2-Methylnaphthalene

Concen: 0.459 ng

RT: 12.180 min Scan# 1174

Delta R.T. -0.004 min

Lab File: BN021573.D

Acq: 02 Sep 2022 22:20




Tgt Ion:142 Resp: 2404

Ion Ratio Lower Upper

142 100

141 14.6 12.2 18.2

115 8.3 7.9 11.9








#13
 Acenaphthene-d10
 Concen: 0.400 ng
 RT: 14.718 min Scan# 14
 Delta R.T. -0.000 min
 Lab File: BN021573.D
 Acq: 02 Sep 2022 22:20




Instrument :
 BNA_N
 ClientSampleId :
 PB147294BS


Tgt Ion:164 Resp: 8353
 Ion Ratio Lower Upper
 164 100
 162 103.5 82.2 123.4
 160 50.6 39.5 59.3



#14
 2,4,6-Tribromophenol
 Concen: 0.372 ng
 RT: 16.208 min Scan# 1608
 Delta R.T. -0.010 min
 Lab File: BN021573.D
 Acq: 02 Sep 2022 22:20

Tgt Ion:330 Resp: 1020
 Ion Ratio Lower Upper
 330 100
 332 93.2 79.3 118.9
 141 44.2 37.1 55.7








#15
2-Fluorobiphenyl
Concen: 0.397 ng
RT: 13.349 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20




Instrument :
BNA_N
ClientSampleId :
PB147294BS

Tgt Ion:172 Resp: 14481
Ion Ratio Lower Upper
172 100
171 34.3 27.8 41.6
170 23.1 18.6 28.0



#16
Acenaphthylene
Concen: 0.381 ng
RT: 14.440 min Scan# 1441
Delta R.T. -0.000 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20

Tgt Ion:152 Resp: 14958
Ion Ratio Lower Upper
152 100
151 19.4 15.4 23.2
153 12.9 10.3 15.5








#17
Acenaphthene
Concen: 0.397 ng
RT: 14.782 min Scan# 1473
Delta R.T. -0.000 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20




Instrument :
BNA_N
ClientSampleId :
PB147294BS


Tgt Ion:154 Resp: 10948
Ion Ratio Lower Upper
154 100
153 113.0 89.5 134.3
152 58.3 45.5 68.3



#18
Fluorene
Concen: 0.399 ng
RT: 15.765 min Scan# 1565
Delta R.T. -0.000 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20

Tgt Ion:166 Resp: 13574
Ion Ratio Lower Upper
166 100
165 99.2 78.7 118.1
167 13.4 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.470 min Scan# 1732
Delta R.T. -0.000 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20




Instrument :
BNA_N
ClientSampleId :
PB147294BS

Tgt Ion:188 Resp: 17321
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.6 9.5 14.3



#21
4-Bromophenyl-phenylether
Concen: 0.386 ng
RT: 16.659 min Scan# 1652
Delta R.T. -0.000 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20

Tgt Ion:248 Resp: 4180
Ion Ratio Lower Upper
248 100
250 102.2 81.8 122.6
141 46.3 38.5 57.7








#22
Hexachlorobenzene
Concen: 0.391 ng
RT: 16.772 min Scan# 1663
Delta R.T. -0.000 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20




Instrument :
BNA_N
ClientSampleId :
PB147294BS


Tgt Ion:284 Resp: 5186
Ion Ratio Lower Upper
284 100
142 41.4 32.3 48.5
249 31.2 25.4 38.2



#23
Atrazine
Concen: 0.366 ng
RT: 16.916 min Scan# 1677
Delta R.T. -0.010 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20

Tgt Ion:200 Resp: 2317
Ion Ratio Lower Upper
200 100
173 33.4 24.3 36.5
215 46.3 38.4 57.6





#24

Pentachlorophenol

Concen: 0.405 ng

RT: 17.121 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN021573.D

Acq: 02 Sep 2022 22:20

Instrument :

BNA_N

ClientSampleId :

PB147294BS





Tgt Ion:266 Resp: 807

Ion Ratio Lower Upper

266 100

264 63.8 49.0 73.6

268 67.8 50.9 76.3



#25

Phenanthrene

Concen: 0.389 ng

RT: 17.511 min Scan# 1735

Delta R.T. -0.000 min

Lab File: BN021573.D

Acq: 02 Sep 2022 22:20




Tgt Ion:178 Resp: 23241

Ion Ratio Lower Upper

178 100

176 19.2 15.4 23.2

179 15.2 12.2 18.4








#26
 Anthracene
 Concen: 0.381 ng
 RT: 17.603 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: BN021573.D
 Acq: 02 Sep 2022 22:20




Instrument :
 BNA_N
 ClientSampleId :
 PB147294BS


Tgt Ion:178 Resp: 18090
 Ion Ratio Lower Upper
 178 100
 176 18.7 14.6 22.0
 179 15.1 12.2 18.2



#27
 Fluoranthene-d10
 Concen: 0.373 ng
 RT: 19.497 min Scan# 2136
 Delta R.T. -0.004 min
 Lab File: BN021573.D
 Acq: 02 Sep 2022 22:20

Tgt Ion:212 Resp: 16623
 Ion Ratio Lower Upper
 212 100
 106 18.1 14.2 21.2
 104 10.0 7.8 11.6








#28
Fluoranthene
Concen: 0.369 ng
RT: 19.527 min Scan# 2143
Delta R.T. -0.004 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20




Instrument :
BNA_N
ClientSampleId :
PB147294BS


Tgt Ion:202 Resp: 22919
Ion Ratio Lower Upper
202 100
101 13.8 10.9 16.3
203 16.9 13.8 20.6



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.655 min Scan# 2441
Delta R.T. -0.009 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20

Tgt Ion:240 Resp: 10971
Ion Ratio Lower Upper
240 100
120 15.9 10.2 15.4#
236 28.3 22.1 33.1









#30
Pyrene
Concen: 0.421 ng
RT: 19.893 min Scan# 21
Delta R.T. -0.004 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20




Instrument :
BNA_N
ClientSampleId :
PB147294BS

Tgt Ion:202 Resp: 24619
Ion Ratio Lower Upper
202 100
200 19.1 13.9 20.9
203 16.2 11.9 17.9



#31
Terphenyl-d14
Concen: 0.445 ng
RT: 20.088 min Scan# 2266
Delta R.T. -0.009 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20

Tgt Ion:244 Resp: 10971
Ion Ratio Lower Upper
244 100
212 9.1 6.5 9.7
122 16.4 9.1 13.7#








#32
Benzo(a)anthracene
Concen: 0.395 ng
RT: 21.637 min Scan# 2439
Delta R.T. -0.009 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20




Instrument :
BNA_N
ClientSampleId :
PB147294BS


Tgt Ion:228 Resp: 15088
Ion Ratio Lower Upper
228 100
226 28.0 22.0 33.0
229 19.7 15.9 23.9



#33
Chrysene
Concen: 0.406 ng
RT: 21.691 min Scan# 2445
Delta R.T. -0.009 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20

Tgt Ion:228 Resp: 18491
Ion Ratio Lower Upper
228 100
226 31.2 24.2 36.2
229 19.6 15.9 23.9









#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.439 ng
 RT: 21.548 min Scan# 2429
 Delta R.T. -0.009 min
 Lab File: BN021573.D
 Acq: 02 Sep 2022 22:20




Instrument :
 BNA_N
 ClientSampleId :
 PB147294BS

Tgt Ion:149 Resp: 6756
 Ion Ratio Lower Upper
 149 100
 167 25.9 21.4 32.0
 279 2.6 2.2 3.2



#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 24.173 min Scan# 3129
 Delta R.T. -0.012 min
 Lab File: BN021573.D
 Acq: 02 Sep 2022 22:20

Tgt Ion:264 Resp: 7463
 Ion Ratio Lower Upper
 264 100
 260 27.0 21.0 31.6
 265 37.5 32.2 48.2








#36
Indeno(1,2,3-cd)pyrene
Concen: 0.396 ng
RT: 26.822 min Scan# 4044
Delta R.T. -0.018 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20




Instrument :
BNA_N
ClientSampleId :
PB147294BS

Tgt Ion:276 Resp: 13344
Ion Ratio Lower Upper
276 100
138 34.0 27.4 41.2
277 24.9 19.9 29.9



#37
Benzo(b)fluoranthene
Concen: 0.390 ng
RT: 23.398 min Scan# 2864
Delta R.T. -0.009 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20

Tgt Ion:252 Resp: 14344
Ion Ratio Lower Upper
252 100
253 24.4 19.8 29.6
125 16.0 13.7 20.5








#38
Benzo(k)fluoranthene
Concen: 0.383 ng
RT: 23.448 min Scan# 21
Delta R.T. -0.012 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20




Instrument :
BNA_N
ClientSampleId :
PB147294BS

Tgt Ion:252	Resp:	14175
Ion Ratio	Lower	Upper
252	100	
253	24.3	19.5 29.3
125	14.9	12.3 18.5



#39
Benzo(a)pyrene
Concen: 0.405 ng
RT: 24.059 min Scan# 3090
Delta R.T. -0.009 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20

Tgt Ion:252	Resp:	10400
Ion Ratio	Lower	Upper
252	100	
253	26.8	21.1 31.7
125	22.4	16.6 25.0









#40
Dibenzo(a,h)anthracene
Concen: 0.383 ng
RT: 26.845 min Scan# 4053
Delta R.T. -0.012 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20

Instrument :
BNA_N
ClientSampleId :
PB147294BS

Tgt Ion:278 Resp: 10389
Ion Ratio Lower Upper
278 100
139 24.3 21.8 32.6
279 26.3 20.4 30.6



#41
Benzo(g,h,i)perylene
Concen: 0.377 ng
RT: 27.638 min Scan# 4314
Delta R.T. -0.015 min
Lab File: BN021573.D
Acq: 02 Sep 2022 22:20

Tgt Ion:276 Resp: 11215
Ion Ratio Lower Upper
276 100
277 24.8 19.7 29.5
138 30.5 23.7 35.5

