

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090319\
 Data File : BN007534.D
 Acq On : 03 Sep 2019 14:37
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 03 16:50:07 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 29 12:18:45 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	197#	-0.02
2	1,4-Dioxane	0.665	0.633	4.8	181#	-0.02
3	n-Nitrosodimethylamine	0.309	0.238	23.0	147	-0.02
4 S	2-Fluorophenol	1.490	1.212	18.7	169#	-0.02
5 S	Phenol-d6	1.953	1.597	18.2	185#	-0.02
6	bis(2-Chloroethyl)ether	1.445	1.495	-3.5	206#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	227#	-0.03
8 S	Nitrobenzene-d5	0.348	0.304	12.6	207#	-0.01
9	Naphthalene	1.143	1.126	1.5	218#	-0.01
10	Hexachlorobutadiene	0.234	0.201	14.1	187#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.665	0.673	-1.2	226#	-0.02
12	2-Methylnaphthalene	0.779	0.797	-2.3	225#	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	249#	-0.02
14 S	2,4,6-Tribromophenol	0.209	0.177	15.3	214#	-0.02
15 S	2-Fluorobiphenyl	1.613	1.583	1.9	239#	0.00
16	Acenaphthylene	2.367	2.173	8.2	226#	-0.02
17	Acenaphthene	1.385	1.333	3.8	235#	-0.02
18	Fluorene	1.751	1.783	-1.8	248#	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	250#	-0.02
20	4-Bromophenyl-phenylether	0.158	0.037	76.6#	67	0.00
21	Hexachlorobenzene	0.268	0.261	2.6	241#	-0.02
22	Pentachlorophenol	0.083	0.074	10.8	238#	-0.02
23	Phenanthrene	1.203	1.325	-10.1	275#	-0.02
24	Anthracene	1.210	1.166	3.6	239#	0.00
25 SURR	Fluoranthene-d10	1.288	1.198	7.0	225#	-0.01
26	Fluoranthene	1.545	1.584	-2.5	250#	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	223#	0.00
28	Pvrene	1.609	1.885	-17.2	255#	0.00
29 S	Terphenyl-d14	1.047	1.091	-4.2	225#	-0.01
30	Benzo(a)anthracene	1.571	1.652	-5.2	230#	-0.01
31	Chrysene	1.516	1.526	-0.7	215#	-0.01
32	Bis(2-ethylhexyl)phthalate	1.177	0.914	22.3	183#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.827	1.331	27.1#	157#	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	181#	-0.01
35	Benzo(b)fluoranthene	1.449	1.583	-9.2	193#	-0.01
36	Benzo(k)fluoranthene	1.432	1.519	-6.1	184#	0.00
37 C	Benzo(a)pyrene	1.379	1.378	0.1	175#	-0.02
38	Dibenzo(a,h)anthracene	1.376	1.197	13.0	152#	-0.02
39	Benzo(a,h,i)perylene	1.371	1.296	5.5	166#	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090319\
Data File : BN007534.D
Acq On : 03 Sep 2019 14:37
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 03 16:50:07 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 29 12:18:45 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			