

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090523\
 Data File : BN027390.D
 Acq On : 06 Sep 2023 00:16
 Operator : MA/JU
 Sample : 04123-05
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BHAA8

Quant Time: Sep 06 01:07:14 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 02 06:16:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.063	152	882	0.400	ng/ul	0.00
4) Naphthalene-d8	10.878	136	2552	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.684	164	1316	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.434	188	3215	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.612	240	3234	0.400	ng/ul	# 0.00
23) Perylene-d12	24.123	264	4079	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.427	96	7137	6.083	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.456	152	1448	0.374	ng/ul	0.00
18) Fluoranthene-d10	19.455	212	3917	0.353	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090523\
Data File : BN027390.D
Acq On : 06 Sep 2023 00:16
Operator : MA/JU
Sample : 04123-05
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BHAA8

Quant Time: Sep 06 01:07:14 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 02 06:16:48 2023
Response via : Initial Calibration

