

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090523\
 Data File : BN027393.D
 Acq On : 06 Sep 2023 02:08
 Operator : MA/JU
 Sample : 04148-06
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BBJW3

Quant Time: Sep 06 02:38:14 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 02 06:16:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.055	152	3872	0.400	ng/ul	0.00
4) Naphthalene-d8	10.878	136	13539	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.684	164	8441	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.435	188	18769	0.400	ng/ul	0.00
17) Chrysene-d12	21.606	240	16306	0.400	ng/ul	0.00
23) Perylene-d12	24.126	264	19222	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.414	96	26472	5.140	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.456	152	7425	0.361	ng/ul	0.00
18) Fluoranthene-d10	19.450	212	21130	0.378	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.452	88	1934	0.371	ng/ul#	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090523\
Data File : BN027393.D
Acq On : 06 Sep 2023 02:08
Operator : MA/JU
Sample : 04148-06
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BBJW3

Quant Time: Sep 06 02:38:14 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 02 06:16:48 2023
Response via : Initial Calibration

