

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090523\
 Data File : BN027386.D
 Acq On : 05 Sep 2023 21:47
 Operator : MA/JU
 Sample : 04123-01
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BH920

Quant Time: Sep 06 00:43:22 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 02 06:16:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.059	152	4294	0.400	ng/ul	0.00
4) Naphthalene-d8	10.883	136	15072	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.689	164	8887	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.434	188	20521	0.400	ng/ul	0.00
17) Chrysene-d12	21.612	240	17231	0.400	ng/ul	0.00
23) Perylene-d12	24.129	264	20742	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.414	96	30673	5.370	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.456	152	7691	0.336	ng/ul	0.00
18) Fluoranthene-d10	19.455	212	21667	0.367	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.452	88	908	0.157	ng/ul#	82


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090523\
Data File : BN027386.D
Acq On : 05 Sep 2023 21:47
Operator : MA/JU
Sample : 04123-01
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BH920

Quant Time: Sep 06 00:43:22 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 02 06:16:48 2023
Response via : Initial Calibration





1,4-Dioxane
Concen: 0.157 ng/ul
RT: 3.452 min Scan# 11
Delta R.T. -0.004 min
Lab File: BN027386.D
Acq: 05 Sep 2023 21:47

Instrument :
BNA_N
ClientSampleId :
BH920

Tgt Ion: 88 Resp: 908
Ion Ratio Lower Upper
88 100
43 45.3 26.8 40.2#
58 38.6 39.9 59.9#

