

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090523\
 Data File : BN027388.D
 Acq On : 05 Sep 2023 23:02
 Operator : MA/JU
 Sample : 04123-03
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BH942

Quant Time: Sep 06 00:43:34 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 02 06:16:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.059	152	5245	0.400	ng/ul	0.00
4) Naphthalene-d8	10.878	136	17704	0.400	ng/ul #	0.00
9) Acenaphthene-d10	14.689	164	10453	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.434	188	23402	0.400	ng/ul	0.00
17) Chrysene-d12	21.606	240	19386	0.400	ng/ul	0.00
23) Perylene-d12	24.126	264	21822	0.400	ng/ul #	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.414	96	30897	4.428	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.456	152	7876	0.293	ng/ul	0.00
18) Fluoranthene-d10	19.450	212	22629	0.341	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.452	88	15756	2.232	ng/ul#	84


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090523\
Data File : BN027388.D
Acq On : 05 Sep 2023 23:02
Operator : MA/JU
Sample : 04123-03
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BH942

Quant Time: Sep 06 00:43:34 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 02 06:16:48 2023
Response via : Initial Calibration





1,4-Dioxane

Concen: 2.232 ng/ul

RT: 3.452 min Scan# 11

Delta R.T. -0.004 min

Lab File: BN027388.D

Acq: 05 Sep 2023 23:02

Instrument :

BNA_N

ClientSampleId :

BH942

Tgt Ion: 88 Resp: 15756

Ion Ratio Lower Upper

88 100

43 29.4 26.8 40.2

58 64.4 39.9 59.9#

