

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090523\
 Data File : BN027392.D
 Acq On : 06 Sep 2023 01:31
 Operator : MA/JU
 Sample : 04148-05
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BBKF1

Quant Time: Sep 06 01:58:16 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 02 06:16:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.055	152	3778	0.400	ng/ul	0.00
4) Naphthalene-d8	10.878	136	13344	0.400	ng/ul #	0.00
9) Acenaphthene-d10	14.685	164	7891	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.434	188	18272	0.400	ng/ul	0.00
17) Chrysene-d12	21.606	240	16027	0.400	ng/ul	0.00
23) Perylene-d12	24.123	264	18636	0.400	ng/ul #	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.414	96	23449	4.666	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.451	152	6970	0.344	ng/ul	0.00
18) Fluoranthene-d10	19.450	212	21954	0.400	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.452	88	1148	0.226	ng/ul#	85


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090523\
Data File : BN027392.D
Acq On : 06 Sep 2023 01:31
Operator : MA/JU
Sample : 04148-05
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BBKF1

Quant Time: Sep 06 01:58:16 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 02 06:16:48 2023
Response via : Initial Calibration





1,4-Dioxane
 Concen: 0.226 ng/ul
 RT: 3.452 min Scan# 11
 Delta R.T. -0.004 min
 Lab File: BN027392.D
 Acq: 06 Sep 2023 01:31

Instrument :
 BNA_N
 ClientSampleId :
 BBKF1

Tgt Ion: 88 Resp: 1148
 Ion Ratio Lower Upper
 88 100
 43 46.5 26.8 40.2#
 58 43.2 39.9 59.9

