

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090619\
 Data File : BN007545.D
 Acq On : 06 Sep 2019 17:00
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 06 18:24:16 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 29 12:18:45 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	154#	-0.02
2	1,4-Dioxane	0.665	0.742	-11.6	166#	0.00
3	n-Nitrosodimethylamine	0.309	0.221	28.5#	107	0.00
4 S	2-Fluorophenol	1.490	1.246	16.4	136	-0.02
5 S	Phenol-d6	1.953	1.473	24.6	134	0.00
6	bis(2-Chloroethyl)ether	1.445	1.370	5.2	147	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	151#	-0.01
8 S	Nitrobenzene-d5	0.348	0.319	8.3	145	0.00
9	Naphthalene	1.143	1.144	-0.1	147	-0.01
10	Hexachlorobutadiene	0.234	0.211	9.8	131	-0.01
11 SURR	2-Methylnaphthalene-d10	0.665	0.616	7.4	138	0.00
12	2-Methylnaphthalene	0.779	0.745	4.4	140	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	146	-0.01
14 S	2,4,6-Tribromophenol	0.209	0.158	24.4	113	-0.01
15 S	2-Fluorobiphenyl	1.613	1.677	-4.0	149	-0.01
16	Acenaphthylene	2.367	1.936	18.2	118	-0.01
17	Acenaphthene	1.385	1.372	0.9	142	-0.01
18	Fluorene	1.751	1.733	1.0	142	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	140	-0.01
20	4-Bromophenyl-phenylether	0.158	0.043	72.8#	44#	0.00
21	Hexachlorobenzene	0.268	0.272	-1.5	141	-0.01
22	Pentachlorophenol	0.083	0.051	38.6#	91	-0.01
23	Phenanthrene	1.203	1.275	-6.0	148	-0.01
24	Anthracene	1.210	1.082	10.6	124	-0.01
25 SURR	Fluoranthene-d10	1.288	1.180	8.4	124	0.00
26	Fluoranthene	1.545	1.505	2.6	133	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	138	-0.01
28	Pvrene	1.609	1.569	2.5	131	0.00
29 S	Terphenyl-d14	1.047	1.042	0.5	133	-0.01
30	Benzo(a)anthracene	1.571	1.352	13.9	117	-0.01
31	Chrysene	1.516	1.604	-5.8	140	-0.01
32	Bis(2-ethylhexyl)phthalate	1.177	0.598	49.2#	74	-0.02
33	Indeno(1,2,3-cd)pyrene	1.827	1.989	-8.9	145	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	142	-0.02
35	Benzo(b)fluoranthene	1.449	1.391	4.0	133	-0.01
36	Benzo(k)fluoranthene	1.432	1.611	-12.5	154#	-0.01
37 C	Benzo(a)pyrene	1.379	1.315	4.6	132	-0.01
38	Dibenzo(a,h)anthracene	1.376	1.411	-2.5	141	-0.02
39	Benzo(a,h,i)perylene	1.371	1.436	-4.7	144	-0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090619\
Data File : BN007545.D
Acq On : 06 Sep 2019 17:00
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 06 18:24:16 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 29 12:18:45 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			