

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090619\
 Data File : BN007557.D
 Acq On : 07 Sep 2019 00:25
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 07 00:58:48 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 29 12:18:45 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	158#	-0.02
2	1,4-Dioxane	0.665	0.711	-6.9	163#	0.00
3	n-Nitrosodimethylamine	0.309	0.193	37.5#	95	0.00
4 S	2-Fluorophenol	1.490	1.206	19.1	135	-0.02
5 S	Phenol-d6	1.953	1.433	26.6#	133	0.00
6	bis(2-Chloroethyl)ether	1.445	1.356	6.2	150	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	157#	-0.03
8 S	Nitrobenzene-d5	0.348	0.308	11.5	145	0.00
9	Naphthalene	1.143	1.133	0.9	151#	-0.01
10	Hexachlorobutadiene	0.234	0.209	10.7	134	-0.01
11 SURR	2-Methylnaphthalene-d10	0.665	0.616	7.4	143	-0.01
12	2-Methylnaphthalene	0.779	0.737	5.4	144	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	152#	-0.02
14 S	2,4,6-Tribromophenol	0.209	0.153	26.8#	113	-0.02
15 S	2-Fluorobiphenyl	1.613	1.636	-1.4	150#	0.00
16	Acenaphthylene	2.367	2.016	14.8	128	-0.02
17	Acenaphthene	1.385	1.373	0.9	147	0.00
18	Fluorene	1.751	1.696	3.1	144	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	150	0.00
20	4-Bromophenyl-phenylether	0.158	0.079	50.0#	87	0.00
21	Hexachlorobenzene	0.268	0.264	1.5	146	-0.02
22	Pentachlorophenol	0.083	0.045	45.8#	87	0.00
23	Phenanthrene	1.203	1.194	0.7	148	-0.02
24	Anthracene	1.210	1.036	14.4	127	0.00
25 SURR	Fluoranthene-d10	1.288	1.145	11.1	129	-0.01
26	Fluoranthene	1.545	1.413	8.5	133	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	144	-0.01
28	Pvrene	1.609	1.519	5.6	133	-0.01
29 S	Terphenyl-d14	1.047	0.975	6.9	130	-0.01
30	Benzo(a)anthracene	1.571	1.389	11.6	125	-0.01
31	Chrysene	1.516	1.521	-0.3	139	-0.01
32	Bis(2-ethylhexyl)phthalate	1.177	0.691	41.3#	89	-0.02
33	Indeno(1,2,3-cd)pyrene	1.827	1.793	1.9	137	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	144	-0.02
35	Benzo(b)fluoranthene	1.449	1.326	8.5	128	-0.02
36	Benzo(k)fluoranthene	1.432	1.564	-9.2	151#	-0.02
37 C	Benzo(a)pyrene	1.379	1.287	6.7	130	-0.02
38	Dibenzo(a,h)anthracene	1.376	1.312	4.7	132	-0.03
39	Benzo(a,h,i)perylene	1.371	1.340	2.3	136	-0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090619\
Data File : BN007557.D
Acq On : 07 Sep 2019 00:25
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 07 00:58:48 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 29 12:18:45 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			