

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090623\
 Data File : BN027429.D
 Acq On : 07 Sep 2023 05:30
 Operator : MA/JU
 Sample : 04123-04
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BH964

Quant Time: Sep 07 06:01:06 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 02 06:16:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.046	152	5101	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.867	136	18158	0.400	ng/ul	-0.02
9) Acenaphthene-d10	14.680	164	11636	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.426	188	26765	0.400	ng/ul	-0.01
17) Chrysene-d12	21.597	240	21400	0.400	ng/ul	-0.01
23) Perylene-d12	24.109	264	21948	0.400	ng/ul	#-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.406	96	32810	4.835	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	12.445	152	9139	0.332	ng/ul	-0.01
18) Fluoranthene-d10	19.446	212	27290	0.372	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090623\
Data File : BN027429.D
Acq On : 07 Sep 2023 05:30
Operator : MA/JU
Sample : 04123-04
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BH964

Quant Time: Sep 07 06:01:06 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 02 06:16:48 2023
Response via : Initial Calibration

