

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090623\
 Data File : BN027454.D
 Acq On : 07 Sep 2023 23:01
 Operator : MA/JU
 Sample : 04123-15
 Misc :
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BH931

Quant Time: Sep 08 01:58:10 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 02 06:16:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.051	152	7883	0.400	ng/ul	0.00
4) Naphthalene-d8	10.872	136	27319	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.680	164	16045	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.430	188	33537	0.400	ng/ul	0.00
17) Chrysene-d12	21.603	240	21975	0.400	ng/ul	0.00
23) Perylene-d12	24.111	264	21854	0.400	ng/ul	#-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.410	96	45366	4.326	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.445	152	14694	0.354	ng/ul	-0.01
18) Fluoranthene-d10	19.446	212	35307	0.469	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090623\
Data File : BN027454.D
Acq On : 07 Sep 2023 23:01
Operator : MA/JU
Sample : 04123-15
Misc :
ALS Vial : 54 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BH931

Quant Time: Sep 08 01:58:10 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 02 06:16:48 2023
Response via : Initial Calibration

