




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090921\
Data File : BN016269.D
Acq On : 09 Sep 2021 22:27
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 10 02:37:40 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 24 17:12:56 2021
Response via : Initial Calibration



TIC: BN016269.D\data.ms

(70) Pentachlorophenol (C)




16.886min (-0.018) 21322.32 ng

response 244305

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.98
264.00	61.60	62.18
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090921\
Data File : BN016269.D
Acq On : 09 Sep 2021 22:27
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 10 02:37:40 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 24 17:12:56 2021
Response via : Initial Calibration



TIC: BN016269.D\data.ms

(77) Benzidine

19.480min (-0.018) 0.00 ng

response 1641515

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.15
183.00	13.20	11.73
0.00	0.00	0.00