

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090921\
 Data File : BN016259.D
 Acq On : 09 Sep 2021 15:43
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN090921

Quant Time: Sep 09 16:15:42 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN090921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 09 15:19:08 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	99	0.00
2	1,4-Dioxane	0.400	0.365	8.8	88	0.00
3	n-Nitrosodimethylamine	0.400	0.428	-7.0	114	0.00
4 S	2-Fluorophenol	0.400	0.404	-1.0	103	0.00
5 S	Phenol-d6	0.400	0.376	6.0	96	0.00
6	bis(2-Chloroethyl)ether	0.400	0.453	-13.2	114	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	100	0.00
8 S	Nitrobenzene-d5	0.400	0.360	10.0	97	0.00
9	Naphthalene	0.400	0.421	-5.2	110	0.00
10	Hexachlorobutadiene	0.400	0.413	-3.2	108	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.418	-4.5	110	0.00
12	2-Methylnaphthalene	0.400	0.432	-8.0	112	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	102	0.00
14 S	2,4,6-Tribromophenol	0.400	0.362	9.5	105	0.00
15 S	2-Fluorobiphenyl	0.400	0.380	5.0	101	0.00
16	Acenaphthylene	0.400	0.421	-5.2	113	0.00
17	Acenaphthene	0.400	0.412	-3.0	110	0.00
18	Fluorene	0.400	0.413	-3.2	111	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	102	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.330	17.5	77	0.00
21	4-Bromophenyl-phenylether	0.400	0.417	-4.2	112	0.00
22	Hexachlorobenzene	0.400	0.424	-6.0	112	0.00
23	Atrazine	0.400	0.400	0.0	116	0.00
24	Pentachlorophenol	0.400	0.381	4.8	106	0.00
25	Phenanthrene	0.400	0.414	-3.5	110	0.01
26	Anthracene	0.400	0.424	-6.0	117	0.00
27 SURR	Fluoranthene-d10	0.400	0.404	-1.0	109	0.00
28	Fluoranthene	0.400	0.416	-4.0	109	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	103	0.00
30	Pyrene	0.400	0.451	-12.7	109	0.00
31 S	Terphenyl-d14	0.400	0.436	-9.0	107	0.00
32	Benzo(a)anthracene	0.400	0.435	-8.7	111	0.01
33	Chrysene	0.400	0.440	-10.0	110	0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.413	-3.2	109	0.01
35 I	Perylene-d12	0.400	0.400	0.0	104	0.01
36	Indeno(1,2,3-cd)pyrene	0.400	0.419	-4.7	115	0.01
37	Benzo(b)fluoranthene	0.400	0.444	-11.0	112	0.00
38	Benzo(k)fluoranthene	0.400	0.463	-15.8	119	0.00
39 C	Benzo(a)pyrene	0.400	0.465	-16.3	123	0.00
40	Dibenzo(a,h)anthracene	0.400	0.430	-7.5	114	0.00
41	Benzo(g,h,i)perylene	0.400	0.416	-4.0	111	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN090921