

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090921\
 Data File : BN016288.D
 Acq On : 10 Sep 2021 13:00
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 10 14:34:52 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN090921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 10 14:34:37 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.142	0.161	-13.4	110	0.00
3	n-Nitrosodimethylamine	0.294	0.261	11.2	95	0.00
4 S	2-Fluorophenol	1.020	1.037	-1.7	104	0.00
5 S	Phenol-d6	1.277	1.290	-1.0	104	0.00
6	bis(2-Chloroethyl)ether	1.053	1.095	-4.0	106	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
8 S	Nitrobenzene-d5	0.321	0.314	2.2	104	0.01
9	Naphthalene	1.069	1.083	-1.3	105	0.00
10	Hexachlorobutadiene	0.218	0.221	-1.4	104	0.00
11 SURR	2-Methylnaphthalene-d10	0.625	0.626	-0.2	104	0.00
12	2-Methylnaphthalene	0.719	0.740	-2.9	105	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	100	-0.01
14 S	2,4,6-Tribromophenol	0.164	0.162	1.2	112	0.00
15 S	2-Fluorobiphenyl	1.566	1.575	-0.6	105	0.00
16	Acenaphthylene	1.790	1.778	0.7	105	0.00
17	Acenaphthene	1.162	1.168	-0.5	105	0.00
18	Fluorene	1.437	1.452	-1.0	106	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
20	4,6-Dinitro-2-methylphenol	0.022	0.006	72.7#	83	0.00
21	4-Bromophenyl-phenylether	0.214	0.216	-0.9	107	0.00
22	Hexachlorobenzene	0.256	0.256	0.0	104	-0.01
23	Atrazine	0.197	0.161	18.3	94	0.00
24	Pentachlorophenol	0.045	0.036	20.0	146	0.00
25	Phenanthrene	1.161	1.171	-0.9	106	0.00
26	Anthracene	1.069	1.070	-0.1	109	0.00
27 SURR	Fluoranthene-d10	1.175	1.185	-0.9	107	0.00
28	Fluoranthene	1.364	1.388	-1.8	106	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
30	Pyrene	1.356	1.620	-19.5	107	0.00
31 S	Terphenyl-d14	1.004	1.174	-16.9	106	0.00
32	Benzo(a)anthracene	1.247	1.364	-9.4	103	0.00
33	Chrysene	1.236	1.338	-8.3	101	0.00
34	Bis(2-ethylhexyl)phthalate	0.768	0.899	-17.1	114	0.00
35 I	Perylene-d12	1.000	1.000	0.0	90	0.00
36	Indeno(1,2,3-cd)pyrene	0.666	0.663	0.5	94	0.00
37	Benzo(b)fluoranthene	1.544	1.723	-11.6	97	0.00
38	Benzo(k)fluoranthene	1.376	1.574	-14.4	101	0.00
39 C	Benzo(a)pyrene	1.148	1.268	-10.5	100	0.00
40	Dibenzo(a,h)anthracene	0.509	0.533	-4.7	96	0.00
41	Benzo(g,h,i)perylene	0.553	0.537	2.9	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4