

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090921\
 Data File : BN016254.D
 Acq On : 09 Sep 2021 12:14
 Operator : CG/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Sep 09 14:07:37 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN090921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 09 14:04:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.864	152	10959	0.400	ng	0.00
7) Naphthalene-d8	10.660	136	53347	0.400	ng	0.00
13) Acenaphthene-d10	14.484	164	30487	0.400	ng	-0.01
19) Phenanthrene-d10	17.237	188	62259	0.400	ng	0.00
29) Chrysene-d12	21.429	240	113377	0.400	ng	# 0.00
35) Perylene-d12	23.817	264	71327	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.457	112	48311	1.562	ng	0.00
5) Phenol-d6	7.034	99	61894	1.565	ng	0.00
8) Nitrobenzene-d5	9.013	82	73063	1.706	ng	0.00
11) 2-Methylnaphthalene-d10	12.247	152	144431	1.700	ng	0.00
14) 2,4,6-Tribromophenol	15.982	330	22894	1.683	ng	0.00
15) 2-Fluorobiphenyl	13.114	172	198967	1.731	ng	0.00
27) Fluoranthene-d10	19.271	212	317420	1.698	ng	0.00
31) Terphenyl-d14	19.867	244	267537	0.934	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.373	88	5644	1.610	ng	97
3) n-Nitrosodimethylamine	3.733	42	14867	1.744	ng	# 100
6) bis(2-Chloroethyl)ether	7.292	93	50527	1.766	ng	99
9) Naphthalene	10.711	128	243587	1.725	ng	100
10) Hexachlorobutadiene	10.990	225	49700	1.729	ng	# 99
12) 2-Methylnaphthalene	12.318	142	166238	1.745	ng	99
16) Acenaphthylene	14.205	152	238958	1.747	ng	100
17) Acenaphthene	14.548	154	152416	1.760	ng	99
18) Fluorene	15.538	166	192262	1.767	ng	100
20) 4,6-Dinitro-2-methylph...	15.626	198	6388	1.447	ng	# 60
21) 4-Bromophenyl-phenylether	16.433	248	57685	1.754	ng	98
22) Hexachlorobenzene	16.555	284	67058	1.728	ng	100
23) Atrazine	16.701	200	56870	1.708	ng	99
24) Pentachlorophenol	16.896	266	12864	1.291	ng	98
25) Phenanthrene	17.273	178	306429	1.758	ng	100
26) Anthracene	17.371	178	293942	1.769	ng	100
28) Fluoranthene	19.301	202	369851	1.789	ng	100
30) Pyrene	19.664	202	376092	1.018	ng	100
32) Benzo(a)anthracene	21.419	228	386205	1.014	ng	98
33) Chrysene	21.461	228	370660	1.002	ng	99
34) Bis(2-ethylhexyl)phtha...	21.334	149	232562	0.907	ng	99
36) Indeno(1,2,3-cd)pyrene	26.291	276	133081	0.671	ng	100
37) Benzo(b)fluoranthene	23.091	252	297653	1.222	ng	98
38) Benzo(k)fluoranthene	23.139	252	269649	1.209	ng	98
39) Benzo(a)pyrene	23.710	252	221220	1.023	ng	97
40) Dibenzo(a,h)anthracene	26.308	278	107540	0.649	ng	95
41) Benzo(g,h,i)perylene	27.048	276	107952	0.648	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090921\
Data File : BN016254.D
Acq On : 09 Sep 2021 12:14
Operator : CG/JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Quant Time: Sep 09 14:07:37 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN090921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 09 14:04:29 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.864 min Scan# 519
 Delta R.T. -0.000 min
 Lab File: BN016254.D
 Acq: 09 Sep 2021 12:14




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion: 152 Resp: 10959
 Ion Ratio Lower Upper
 152 100
 150 156.4 125.6 188.4
 115 62.0 50.2 75.2



#2
 1,4-Dioxane
 Concen: 1.610 ng
 RT: 3.373 min Scan# 32
 Delta R.T. -0.000 min
 Lab File: BN016254.D
 Acq: 09 Sep 2021 12:14

Tgt Ion: 88 Resp: 5644
 Ion Ratio Lower Upper
 88 100
 43 29.6 20.3 30.5
 58 86.1 69.9 104.9









#3
n-Nitrosodimethylamine
Concen: 1.744 ng
RT: 3.733 min Scan# 71
Delta R.T. -0.009 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 42 Resp: 14867
Ion Ratio Lower Upper
42 100
74 141.8 113.6 170.4
44 14.7 0.0 0.0#



#4
2-Fluorophenol
Concen: 1.562 ng
RT: 5.457 min Scan# 258
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14

Tgt Ion: 112 Resp: 48311
Ion Ratio Lower Upper
112 100
64 58.7 47.3 70.9
63 28.4 23.4 35.0









#5
Phenol-d6
Concen: 1.565 ng
RT: 7.034 min Scan# 429
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	99	Resp:	61894
Ion	Ratio	Lower	Upper
99	100		
42	17.2	14.0	21.0
71	29.7	23.8	35.6



#6
bis(2-Chloroethyl)ether
Concen: 1.766 ng
RT: 7.292 min Scan# 457
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14


Tgt Ion:	93	Resp:	50527
Ion	Ratio	Lower	Upper
93	100		
63	77.8	63.4	95.0
95	32.9	26.8	40.2








#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.660 min Scan# 753
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6







Tgt Ion:	136	Resp:	53347
Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.8	13.2
54	6.2	5.3	7.9
68	4.7	3.9	5.9



#8
Nitrobenzene-d5
Concen: 1.706 ng
RT: 9.013 min Scan# 623
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14

Tgt Ion:	82	Resp:	73063
Ion	Ratio	Lower	Upper
82	100		
128	41.4	31.7	47.5
54	56.2	46.4	69.6









#9
Naphthalene
Concen: 1.725 ng
RT: 10.711 min Scan# 757
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:128 Resp: 243587
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 12.6 10.1 15.1



#10
Hexachlorobutadiene
Concen: 1.729 ng
RT: 10.990 min Scan# 779
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14

Tgt Ion:225 Resp: 49700
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.8 51.5 77.3









#11
2-Methylnaphthalene-d10
Concen: 1.700 ng
RT: 12.247 min Scan# 986
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 144431
Ion Ratio Lower Upper
152 100
151 20.1 16.1 24.1



#12
2-Methylnaphthalene
Concen: 1.745 ng
RT: 12.318 min Scan# 1002
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14

Tgt Ion:142 Resp: 166238
Ion Ratio Lower Upper
142 100
141 88.3 69.8 104.8
115 30.6 24.4 36.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.484 min Scan# 1200
Delta R.T. -0.013 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:164 Resp: 30487
Ion Ratio Lower Upper
164 100
162 103.0 79.0 118.4
160 46.5 34.1 51.1



#14
2,4,6-Tribromophenol
Concen: 1.683 ng
RT: 15.982 min Scan# 1318
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14

Tgt Ion:330 Resp: 22894
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 31.1 25.0 37.4









#15
2-Fluorobiphenyl
Concen: 1.731 ng
RT: 13.114 min Scan# 1092
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:172 Resp: 198967
Ion Ratio Lower Upper
172 100
171 35.0 27.8 41.6
170 22.7 18.2 27.2



#16
Acenaphthylene
Concen: 1.747 ng
RT: 14.205 min Scan# 1178
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14

Tgt Ion:152 Resp: 238958
Ion Ratio Lower Upper
152 100
151 19.0 15.2 22.8
153 13.1 10.5 15.7









#17
Acenaphthene
Concen: 1.760 ng
RT: 14.548 min Scan# 1205
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 152416
Ion Ratio Lower Upper
154 100
153 110.4 89.3 133.9
152 51.4 42.2 63.4



#18
Fluorene
Concen: 1.767 ng
RT: 15.538 min Scan# 1283
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14

Tgt Ion:166 Resp: 192262
Ion Ratio Lower Upper
166 100
165 95.9 76.7 115.1
167 13.4 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.237 min Scan# 1421
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:188 Resp: 62259
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 8.7 7.0 10.6



#20
4,6-Dinitro-2-methylphenol
Concen: 1.447 ng
RT: 15.626 min Scan# 1290
Delta R.T. -0.013 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14

Tgt Ion:198 Resp: 6388
Ion Ratio Lower Upper
198 100
182 10.3 26.6 39.8#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 1.754 ng
RT: 16.433 min Scan# 1355
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	248	Resp:	57685
Ion Ratio	Lower	Upper	
248	100		
250	101.3	80.5	120.7
141	39.3	34.8	52.2



#22
Hexachlorobenzene
Concen: 1.728 ng
RT: 16.555 min Scan# 1365
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14

Tgt Ion:	284	Resp:	67058
Ion Ratio	Lower	Upper	
284	100		
142	28.5	22.6	34.0
249	30.5	24.4	36.6









#23
Atrazine
Concen: 1.708 ng
RT: 16.701 min Scan# 1377
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:200 Resp: 56870
Ion Ratio Lower Upper
200 100
173 22.1 18.7 28.1
215 51.9 41.6 62.4



#24
Pentachlorophenol
Concen: 1.291 ng
RT: 16.896 min Scan# 1393
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14

Tgt Ion:266 Resp: 12864
Ion Ratio Lower Upper
266 100
264 62.5 50.6 75.8
268 64.1 52.6 79.0









#25
Phenanthrene
Concen: 1.758 ng
RT: 17.273 min Scan# 1424
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:178 Resp: 306429
Ion Ratio Lower Upper
178 100
176 18.0 14.4 21.6
179 15.3 12.2 18.4



#26
Anthracene
Concen: 1.769 ng
RT: 17.371 min Scan# 1432
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14

Tgt Ion:178 Resp: 293942
Ion Ratio Lower Upper
178 100
176 17.4 13.9 20.9
179 15.4 12.2 18.4









#27
Fluoranthene-d10
Concen: 1.698 ng
RT: 19.271 min Scan# 1738
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:212 Resp: 317420
Ion Ratio Lower Upper
212 100
106 12.1 9.8 14.6
104 6.7 5.4 8.2



#28
Fluoranthene
Concen: 1.789 ng
RT: 19.301 min Scan# 1744
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14

Tgt Ion:202 Resp: 369851
Ion Ratio Lower Upper
202 100
101 10.6 8.5 12.7
203 17.4 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.429 min Scan# 2020
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:240 Resp: 113377
Ion Ratio Lower Upper
240 100
120 8.7 5.3 7.9#
236 25.3 19.8 29.8



#30
Pyrene
Concen: 1.018 ng
RT: 19.664 min Scan# 1817
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14

Tgt Ion:202 Resp: 376092
Ion Ratio Lower Upper
202 100
200 19.7 15.8 23.6
203 17.6 13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.934 ng
 RT: 19.867 min Scan# 1858
 Delta R.T. -0.000 min
 Lab File: BN016254.D
 Acq: 09 Sep 2021 12:14




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:244	Resp: 267537
Ion Ratio	Lower Upper
244 100	
212 6.8	5.5 8.3
122 11.2	9.0 13.4



#32
 Benzo(a)anthracene
 Concen: 1.014 ng
 RT: 21.419 min Scan# 2019
 Delta R.T. 0.011 min
 Lab File: BN016254.D
 Acq: 09 Sep 2021 12:14

Tgt Ion:228	Resp: 386205
Ion Ratio	Lower Upper
228 100	
226 24.7	20.8 31.2
229 20.0	15.8 23.6









#33
Chrysene
Concen: 1.002 ng
RT: 21.461 min Scan# 2023
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:228 Resp: 370660
Ion Ratio Lower Upper
228 100
226 29.1 23.0 34.4
229 19.1 15.6 23.4



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.907 ng
RT: 21.334 min Scan# 2011
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14

Tgt Ion:149 Resp: 232562
Ion Ratio Lower Upper
149 100
167 29.8 23.4 35.0
279 4.8 3.9 5.9





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.817 min Scan# 2524

Delta R.T. 0.003 min

Lab File: BN016254.D

Acq: 09 Sep 2021 12:14





Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6

Tgt Ion:264	Resp:	71327
Ion Ratio	Lower	Upper
264	100	
260	23.6	18.9 28.3
265	24.8	23.7 35.5



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.671 ng




RT: 26.291 min Scan# 3247


Delta R.T. 0.003 min

Lab File: BN016254.D

Acq: 09 Sep 2021 12:14

Tgt Ion:276	Resp:	133081
Ion Ratio	Lower	Upper
276	100	
138	23.9	19.3 28.9
277	25.1	20.2 30.2









#37
Benzo(b)fluoranthene
Concen: 1.222 ng
RT: 23.091 min Scan# 2312
Delta R.T. 0.003 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 297653
Ion Ratio Lower Upper
252 100
253 22.0 18.2 27.2
125 9.6 8.4 12.6



#38
Benzo(k)fluoranthene
Concen: 1.209 ng
RT: 23.139 min Scan# 2326
Delta R.T. 0.003 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14


Tgt Ion:252 Resp: 269649
Ion Ratio Lower Upper
252 100
253 22.0 18.2 27.2
125 9.5 8.2 12.4








#39
Benzo(a)pyrene
Concen: 1.023 ng
RT: 23.710 min Scan# 2493
Delta R.T. 0.003 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6







Tgt Ion:252 Resp: 221220
Ion Ratio Lower Upper
252 100
253 22.0 18.7 28.1
125 11.0 9.9 14.9



#40
Dibenzo(a,h)anthracene
Concen: 0.649 ng
RT: 26.308 min Scan# 3252
Delta R.T. -0.000 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14

Tgt Ion:278 Resp: 107540
Ion Ratio Lower Upper
278 100
139 15.6 14.2 21.2
279 24.0 21.3 31.9





#41
Benzo(g,h,i)perylene
Concen: 0.648 ng
RT: 27.048 min Scan# 3468
Delta R.T. 0.003 min
Lab File: BN016254.D
Acq: 09 Sep 2021 12:14

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:276 Resp: 107952
Ion Ratio Lower Upper
276 100
277 24.1 20.5 30.7
138 20.8 18.2 27.4

