

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090921\
 Data File : BN016286.D
 Acq On : 10 Sep 2021 09:58
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Manual Integrations
 APPROVED

mohammad
 9/10/2021 3:24:14 PM

Quant Time: Sep 10 11:03:53 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN090921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 09 15:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.864	152	12954	0.400	ng	0.00
7) Naphthalene-d8	10.660	136	66728	0.400	ng	0.00
13) Acenaphthene-d10	14.497	164	38874	0.400	ng	0.00
19) Phenanthrene-d10	17.236	188	79727	0.400	ng	0.00
29) Chrysene-d12	21.429	240	71612	0.400	ng	# 0.00
35) Perylene-d12	23.820	264	46480	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.457	112	12963	0.392	ng	0.00
5) Phenol-d6	7.052	99	16377	0.396	ng	0.02
8) Nitrobenzene-d5	9.025	82	20179	0.376	ng	0.01
11) 2-Methylnaphthalene-d10	12.247	152	40542	0.389	ng	0.00
14) 2,4,6-Tribromophenol	15.981	330	6849	0.429	ng	0.00
15) 2-Fluorobiphenyl	13.114	172	57197	0.376	ng	0.00
27) Fluoranthene-d10	19.271	212	91321	0.390	ng	0.00
31) Terphenyl-d14	19.867	244	78355	0.436	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.373	88	1441	0.313	ng	# 84
3) n-Nitrosodimethylamine	3.742	42	3379	0.354	ng	# 98
6) bis(2-Chloroethyl)ether	7.292	93	13739	0.403	ng	99
9) Naphthalene	10.711	128	68815	0.386	ng	100
10) Hexachlorobutadiene	10.990	225	14041	0.386	ng	# 99
12) 2-Methylnaphthalene	12.318	142	47195	0.393	ng	99
16) Acenaphthylene	14.205	152	66989	0.385	ng	100
17) Acenaphthene	14.548	154	43066	0.381	ng	100
18) Fluorene	15.537	166	54692	0.391	ng	99
20) 4,6-Dinitro-2-methylph...	15.626	198	913	0.464	ng	92
21) 4-Bromophenyl-phenylether	16.433	248	16465	0.386	ng	98
22) Hexachlorobenzene	16.543	284	19216	0.376	ng	100
23) Atrazine	16.701	200	15221	0.388	ng	96
24) Pentachlorophenol	16.896	266	3703	0.568	ng	97
25) Phenanthrene	17.273	178	87909	0.380	ng	99
26) Anthracene	17.370	178	83999	0.394	ng	100
28) Fluoranthene	19.301	202	105004	0.386	ng	99
30) Pyrene	19.663	202	106875	0.440	ng	100
32) Benzo(a)anthracene	21.418	228	94648	0.424	ng	98
33) Chrysene	21.461	228	91308	0.413	ng	98
34) Bis(2-ethylhexyl)phtha...	21.344	149	66088	0.481	ng	100
36) Indeno(1,2,3-cd)pyrene	26.288	276	33220	0.429	ng	99
37) Benzo(b)fluoranthene	23.091	252	74157m	0.413	ng	
38) Benzo(k)fluoranthene	23.139	252	66939	0.419	ng	# 96
39) Benzo(a)pyrene	23.714	252	56223	0.421	ng	# 87
40) Dibenzo(a,h)anthracene	26.308	278	27515	0.465	ng	# 92
41) Benzo(g,h,i)perylene	27.048	276	26135	0.407	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090921\
Data File : BN016286.D
Acq On : 10 Sep 2021 09:58
Operator : CG/JU
Sample : SSTCCCC0.4
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTCCCC0.4EC

Manual Integrations
APPROVED


mohammad
9/10/2021 3:24:14 PM

Quant Time: Sep 10 11:03:53 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN090921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 09 15:19:08 2021
Response via : Initial Calibration




mohammad



9/10/2021 3:24:14 PM



#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.864 min Scan# 519
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58



Tgt Ion: 152 Resp: 12954
Ion Ratio Lower Upper
152 100
150 154.7 125.6 188.4
115 61.8 50.2 75.2




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED


mohammad
9/10/2021 3:24:14 PM

mohammad



9/10/2021 3:24:14 PM




#2
1,4-Dioxane
Concen: 0.313 ng
RT: 3.373 min Scan# 32
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58



Tgt Ion: 88 Resp: 1441
Ion Ratio Lower Upper
88 100
43 22.3 20.3 30.5
58 69.6 69.9 104.9#





#3
n-Nitrosodimethylamine
Concen: 0.354 ng
RT: 3.742 min Scan# 72
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58




Tgt Ion: 42 Resp: 3379

Ion	Ratio	Lower	Upper
42	100		
74	144.5	113.6	170.4
44	0.0	0.0	0.0

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Manual Integrations
APPROVED

mohammad
9/10/2021 3:24:14 PM



mohammad




9/10/2021 3:24:14 PM




#4
2-Fluorophenol
Concen: 0.392 ng
RT: 5.457 min Scan# 258
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58

Tgt Ion: 112 Resp: 12963

Ion	Ratio	Lower	Upper
112	100		
64	59.0	47.3	70.9
63	28.4	23.4	35.0








#5
Phenol-d6
Concen: 0.396 ng
RT: 7.052 min Scan# 431
Delta R.T. 0.018 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58

Tgt Ion: 99 Resp: 16377
Ion Ratio Lower Upper
99 100
42 17.6 14.0 21.0
71 30.0 23.8 35.6

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Manual Integrations
APPROVED

mohammad
9/10/2021 3:24:14 PM






mohammad


9/10/2021 3:24:14 PM



#6
bis(2-Chloroethyl)ether
Concen: 0.403 ng
RT: 7.292 min Scan# 457
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58


Tgt Ion: 93 Resp: 13739
Ion Ratio Lower Upper
93 100
63 78.5 63.4 95.0
95 32.9 26.8 40.2





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.660 min Scan# 753
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:	136	Resp:	66728
Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.8	13.2
54	6.3	5.3	7.9
68	4.7	3.9	5.9


Manual Integrations
APPROVED

mohammad
9/10/2021 3:24:14 PM




mohammad



9/10/2021 3:24:14 PM




#8
Nitrobenzene-d5
Concen: 0.376 ng
RT: 9.025 min Scan# 624
Delta R.T. 0.013 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58



Tgt Ion:	82	Resp:	20179
Ion	Ratio	Lower	Upper
82	100		
128	53.3	31.7	47.5#
54	48.1	46.4	69.6





#9
Naphthalene
Concen: 0.386 ng
RT: 10.711 min Scan# 757
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58





Tgt Ion:	128	Resp:	68815
Ion Ratio	Lower	Upper	
128	100		
129	11.2	8.9	13.3
127	12.6	10.1	15.1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED




mohammad
9/10/2021 3:24:14 PM


9/10/2021 3:24:14 PM



#10
Hexachlorobutadiene
Concen: 0.386 ng
RT: 10.990 min Scan# 779
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58

Tgt Ion:	225	Resp:	14041
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.7	51.5	77.3





#11
2-Methylnaphthalene-d10
Concen: 0.389 ng
RT: 12.247 min Scan# 986
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58





Tgt Ion:152 Resp: 40542
Ion Ratio Lower Upper
152 100
151 20.2 16.1 24.1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED




mohammad
9/10/2021 3:24:14 PM


9/10/2021 3:24:14 PM



#12
2-Methylnaphthalene
Concen: 0.393 ng
RT: 12.318 min Scan# 1002
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58


Tgt Ion:142 Resp: 47195
Ion Ratio Lower Upper
142 100
141 87.9 69.8 104.8
115 30.9 24.4 36.6





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.497 min Scan# 1201
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:164 Resp: 38874
Ion Ratio Lower Upper
164 100
162 97.9 79.0 118.4
160 41.6 34.1 51.1

Manual Integrations
APPROVED



mohammad
9/10/2021 3:24:14 PM




#14
2,4,6-Tribromophenol
Concen: 0.429 ng
RT: 15.981 min Scan# 1318
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58



Tgt Ion:330 Resp: 6849
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 31.0 25.0 37.4





#15
2-Fluorobiphenyl
Concen: 0.376 ng
RT: 13.114 min Scan# 1092
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58





Tgt Ion:	172	Resp:	57197
Ion Ratio	Lower	Upper	
172	100		
171	34.9	27.8	41.6
170	23.0	18.2	27.2

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED




mohammad
9/10/2021 3:24:14 PM


9/10/2021 3:24:14 PM



#16
Acenaphthylene
Concen: 0.385 ng
RT: 14.205 min Scan# 1178
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58


Tgt Ion:	152	Resp:	66989
Ion Ratio	Lower	Upper	
152	100		
151	18.9	15.2	22.8
153	13.0	10.5	15.7





#17
Acenaphthene
Concen: 0.381 ng
RT: 14.548 min Scan# 1205
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:154 Resp: 43066
Ion Ratio Lower Upper
154 100
153 111.3 89.3 133.9
152 52.6 42.2 63.4

Manual Integrations
APPROVED



mohammad
9/10/2021 3:24:14 PM




#18
Fluorene
Concen: 0.391 ng
RT: 15.537 min Scan# 1283
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58



Tgt Ion:166 Resp: 54692
Ion Ratio Lower Upper
166 100
165 96.7 76.7 115.1
167 13.6 10.9 16.3





#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.236 min Scan# 1421
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58





Tgt Ion:	188	Resp:	79727
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	8.9	7.0	10.6

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED




mohammad
9/10/2021 3:24:14 PM


9/10/2021 3:24:14 PM



#20
4,6-Dinitro-2-methylphenol
Concen: 0.464 ng
RT: 15.626 min Scan# 1290
Delta R.T. -0.013 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58


Tgt Ion:	198	Resp:	913
Ion Ratio	Lower	Upper	
198	100		
182	28.9	26.6	39.8
77	0.0	0.0	0.0





#21
4-Bromophenyl-phenylether
Concen: 0.386 ng
RT: 16.433 min Scan# 1355
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:	248	Resp:	16465
Ion Ratio	Lower	Upper	
248	100		
250	101.9	80.5	120.7
141	41.8	34.8	52.2

Manual Integrations
APPROVED



mohammad
9/10/2021 3:24:14 PM




#22
Hexachlorobenzene
Concen: 0.376 ng
RT: 16.543 min Scan# 1364
Delta R.T. -0.012 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58




Tgt Ion:	284	Resp:	19216
Ion Ratio	Lower	Upper	
284	100		
142	28.5	22.6	34.0
249	30.3	24.4	36.6









#23
Atrazine
Concen: 0.388 ng
RT: 16.701 min Scan# 1377
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58





Tgt Ion	Ratio	Lower	Upper
200	100		
173	27.9	18.7	28.1
215	50.7	41.6	62.4



#24
Pentachlorophenol
Concen: 0.568 ng
RT: 16.896 min Scan# 1393
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58



Tgt Ion	Ratio	Lower	Upper
266	100		
264	61.4	50.6	75.8
268	63.6	52.6	79.0




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED

mohammad
9/10/2021 3:24:14 PM

9/10/2021 3:24:14 PM



#25
Phenanthrene
Concen: 0.380 ng
RT: 17.273 min Scan# 1424
Delta R.T. 0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58





Tgt Ion:	178	Resp:	87909
Ion Ratio	Lower	Upper	
178	100		
176	18.2	14.4	21.6
179	15.5	12.2	18.4

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED




mohammad
9/10/2021 3:24:14 PM


9/10/2021 3:24:14 PM



#26
Anthracene
Concen: 0.394 ng
RT: 17.370 min Scan# 1432
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58


Tgt Ion:	178	Resp:	83999
Ion Ratio	Lower	Upper	
178	100		
176	17.4	13.9	20.9
179	15.2	12.2	18.4





#27
Fluoranthene-d10
Concen: 0.390 ng
RT: 19.271 min Scan# 1738
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC






Tgt Ion:212 Resp: 91321
Ion Ratio Lower Upper
212 100
106 12.0 9.8 14.6
104 6.7 5.4 8.2


Manual Integrations
APPROVED

mohammad
9/10/2021 3:24:14 PM



9/10/2021 3:24:14 PM




#28
Fluoranthene
Concen: 0.386 ng
RT: 19.301 min Scan# 1744
Delta R.T. 0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58



Tgt Ion:202 Resp: 105004
Ion Ratio Lower Upper
202 100
101 10.1 8.5 12.7
203 17.3 13.8 20.8





#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.429 min Scan# 2020
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58





Tgt Ion	Ratio	Lower	Upper
240	100		
120	10.0	5.3	7.9#
236	25.4	19.8	29.8

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED




mohammad
9/10/2021 3:24:14 PM


9/10/2021 3:24:14 PM



#30
Pyrene
Concen: 0.440 ng
RT: 19.663 min Scan# 1817
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58

Tgt Ion	Ratio	Lower	Upper
202	100		
200	19.8	15.8	23.6
203	17.7	13.9	20.9





#31
Terphenyl-d14
Concen: 0.436 ng
RT: 19.867 min Scan# 1858
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58





Tgt Ion:244 Resp: 78355
Ion Ratio Lower Upper
244 100
212 7.3 5.5 8.3
122 11.9 9.0 13.4

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED




mohammad
9/10/2021 3:24:14 PM


9/10/2021 3:24:14 PM




#32
Benzo(a)anthracene
Concen: 0.424 ng
RT: 21.418 min Scan# 2019
Delta R.T. 0.010 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58

Tgt Ion:228 Resp: 94648
Ion Ratio Lower Upper
228 100
226 25.3 20.8 31.2
229 21.0 15.8 23.6







#33
Chrysene
Concen: 0.413 ng
RT: 21.461 min Scan# 2023
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58



Tgt Ion: 228 Resp: 91308
Ion Ratio Lower Upper
228 100
226 30.0 23.0 34.4
229 20.3 15.6 23.4




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Manual Integrations
APPROVED

mohammad
9/10/2021 3:24:14 PM



9/10/2021 3:24:14 PM




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.481 ng
RT: 21.344 min Scan# 2012
Delta R.T. 0.010 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58




Tgt Ion: 149 Resp: 66088
Ion Ratio Lower Upper
149 100
167 29.2 23.4 35.0
279 4.8 3.9 5.9





#35
Perylene-d12
Concen: 0.400 ng
RT: 23.820 min Scan# 2525
Delta R.T. 0.007 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC






Tgt Ion:264 Resp: 46480
Ion Ratio Lower Upper
264 100
260 24.9 18.9 28.3
265 37.4 23.7 35.5#

Manual Integrations
APPROVED




mohammad
9/10/2021 3:24:14 PM


9/10/2021 3:24:14 PM




#36
Indeno(1,2,3-cd)pyrene
Concen: 0.429 ng
RT: 26.288 min Scan# 3246
Delta R.T. -0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58

Tgt Ion:276 Resp: 33220
Ion Ratio Lower Upper
276 100
138 24.9 19.3 28.9
277 24.9 20.2 30.2









#37
Benzo(b)fluoranthene
Concen: 0.413 ng m
RT: 23.091 min Scan# 2312
Delta R.T. 0.003 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58





Tgt Ion:252 Resp: 74157
Ion Ratio Lower Upper
252 100
253 25.0 18.2 27.2
125 14.9 8.4 12.6#



#38
Benzo(k)fluoranthene
Concen: 0.419 ng
RT: 23.139 min Scan# 2326
Delta R.T. 0.003 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58



Tgt Ion:252 Resp: 66939
Ion Ratio Lower Upper
252 100
253 24.4 18.2 27.2
125 12.5 8.2 12.4#




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Manual Integrations
APPROVED

mohammad
9/10/2021 3:24:14 PM



9/10/2021 3:24:14 PM



#39
Benzo(a)pyrene
Concen: 0.421 ng
RT: 23.714 min Scan# 2494
Delta R.T. 0.007 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58



Tgt Ion:252 Resp: 56223
Ion Ratio Lower Upper
252 100
253 27.1 18.7 28.1
125 21.8 9.9 14.9#




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Manual Integrations
APPROVED

mohammad
9/10/2021 3:24:14 PM



9/10/2021 3:24:14 PM




#40
Dibenzo(a,h)anthracene
Concen: 0.465 ng
RT: 26.308 min Scan# 3252
Delta R.T. 0.000 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58




Tgt Ion:278 Resp: 27515
Ion Ratio Lower Upper
278 100
139 24.4 14.2 21.2#
279 28.5 21.3 31.9







#41
Benzo(g,h,i)perylene
Concen: 0.407 ng
RT: 27.048 min Scan# 3468
Delta R.T. 0.003 min
Lab File: BN016286.D
Acq: 10 Sep 2021 09:58



Tgt Ion:	276	Resp:	26135
Ion Ratio	Lower	Upper	
276	100		
277	27.1	20.5	30.7
138	26.6	18.2	27.4



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED

mohammad
9/10/2021 3:24:14 PM

9/10/2021 3:24:14 PM