

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090923\
 Data File : BN027497.D
 Acq On : 09 Sep 2023 12:15
 Operator : MA/JU
 Sample : PB155239BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK239

Quant Time: Sep 10 00:09:53 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 08 03:55:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.042	152	13632	0.400	ng/u1	0.00
4) Naphthalene-d8	10.861	136	47640	0.400	ng/u1	-0.01
9) Acenaphthene-d10	14.675	164	29752	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.421	188	65244	0.400	ng/u1	0.00
17) Chrysene-d12	21.597	240	49177	0.400	ng/u1	# 0.00
23) Perylene-d12	24.102	264	47809	0.400	ng/u1	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.406	96	82479	4.548	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.440	152	21535	0.298	ng/u1	0.00
18) Fluoranthene-d10	19.441	212	52298	0.310	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090923\
Data File : BN027497.D
Acq On : 09 Sep 2023 12:15
Operator : MA/JU
Sample : PB155239BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK239

Quant Time: Sep 10 00:09:53 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 08 03:55:54 2023
Response via : Initial Calibration

