

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090923\
 Data File : BN027529.D
 Acq On : 10 Sep 2023 08:04
 Operator : MA/JU
 Sample : PB155304BL
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK304

Quant Time: Sep 11 01:18:08 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 08 03:55:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.038	152	7212	0.400	ng/ul	0.00
4) Naphthalene-d8	10.861	136	24575	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.670	164	15005	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.417	188	32951	0.400	ng/ul	0.00
17) Chrysene-d12	21.595	240	23631	0.400	ng/ul	0.00
23) Perylene-d12	24.097	264	22059	0.400	ng/ul	#-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.405	96	47252	4.925	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.440	152	10789	0.289	ng/ul	0.00
18) Fluoranthene-d10	19.436	212	25327	0.313	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090923\
Data File : BN027529.D
Acq On : 10 Sep 2023 08:04
Operator : MA/JU
Sample : PB155304BL
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK304

Quant Time: Sep 11 01:18:08 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 08 03:55:54 2023
Response via : Initial Calibration

