

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090923\
 Data File : BN027552.D
 Acq On : 10 Sep 2023 22:11
 Operator : MA/JU
 Sample : 04262-06
 Misc :
 ALS Vial : 61 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BBHM8

Quant Time: Sep 11 02:13:49 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 08 03:55:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.034	152	7446	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.856	136	25744	0.400	ng/ul	-0.02
9) Acenaphthene-d10	14.666	164	15361	0.400	ng/ul	-0.01
13) Phenanthrene-d10	17.413	188	33396	0.400	ng/ul	-0.01
17) Chrysene-d12	21.589	240	23812	0.400	ng/ul	0.00
23) Perylene-d12	24.094	264	20896	0.400	ng/ul	#-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.406	96	41013	4.141	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.434	152	11011	0.282	ng/ul	-0.01
18) Fluoranthene-d10	19.436	212	29903	0.367	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.443	88	363	0.036	ng/ul#	55

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090923\
Data File : BN027552.D
Acq On : 10 Sep 2023 22:11
Operator : MA/JU
Sample : 04262-06
Misc :
ALS Vial : 61 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BBHM8

Quant Time: Sep 11 02:13:49 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 08 03:55:54 2023
Response via : Initial Calibration

