



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090923\
Data File : BN027494.D
Acq On : 09 Sep 2023 10:25
Operator : MA/JU
Sample : DFTPP372
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP372

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri Sep 08 03:55:54 2023



AutoFind: Scans 2532, 2533, 2534; Background Corrected with Scan 2525

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	27.3	205457	PASS
68	69	0.00	2	1.8	4454	PASS
69	198	0.00	100	33.0	248386	PASS
70	69	0.00	2	0.6	1425	PASS
127	198	10	80	43.4	326763	PASS
197	198	0.00	2	0.4	2708	PASS
198	198	100	100	100.0	752853	PASS
199	198	5	9	6.6	49832	PASS
275	198	10	60	27.6	208021	PASS
365	198	1	100	3.5	26019	PASS
441	198	0.01	100	13.1	98861	PASS
442	198	50	100	85.8	645867	PASS
443	442	15	24	18.5	119789	PASS

m/z	Abundance
39.00	358.0
39.90	843.0
41.00	506.0
42.00	157.0
42.90	556.0
43.90	1424.0
44.90	317.0
48.90	343.0
51.00	282.0
55.00	491.0
57.00	253.0
65.00	164.0
67.00	269.0
69.00	275.0
71.00	164.0
73.00	1691.0
74.00	202.0
75.00	179.0
76.90	717.0
77.90	201.0
79.00	308.0
81.00	297.0
82.90	215.0
83.90	446.0
85.80	224.0
87.00	155.0
91.00	336.0
93.00	292.0
94.90	226.0
96.00	367.0
96.50	185.0
97.10	212.0
102.90	188.0
104.90	209.0
107.00	223.0
114.90	213.0
117.00	197.0
118.90	291.0
121.00	188.0
128.00	164.0
129.00	188.0
132.90	450.0
135.00	357.0
146.90	536.0
149.10	179.0
165.00	164.0
177.00	220.0
191.00	526.0
191.90	196.0
192.90	533.0
207.00	3611.0
208.00	680.0
208.90	449.0
221.10	193.0
266.90	175.0
281.00	815.0
281.90	278.0
283.00	226.0
355.10	174.0

Instrument :
BNA_N
ClientSampleId :
DFTPP372

m/z	Abundance
36.00	207.0
37.00	1357.0
38.00	3147.0
39.00	17576.0
40.00	1408.0
41.00	825.0
42.00	185.0
43.00	474.0
44.00	1563.0
45.00	727.0
48.00	196.0
49.00	1808.0
50.00	60024.0
51.00	227328.0
52.00	11758.0
53.00	618.0
55.00	1393.0
55.90	7022.0
57.00	14915.0
57.90	747.0
59.00	423.0
60.00	359.0
61.00	3190.0
62.00	3356.0
63.00	10043.0
64.00	1342.0
65.00	4925.0
65.90	424.0
67.00	723.0
68.00	4907.0
69.00	268992.0
70.00	1541.0
71.00	179.0
72.00	190.0
73.00	8310.0
74.00	29832.0
75.00	45176.0
76.00	16004.0
77.00	320000.0
78.00	23200.0
79.00	21368.0
80.00	15626.0
81.00	22632.0
82.00	5648.0
83.00	5260.0
84.00	832.0
85.00	3739.0
86.00	6816.0
87.00	3100.0
88.00	1313.0
88.90	694.0
90.00	166.0
91.00	5131.0
92.00	5871.0
93.00	37120.0
94.00	2486.0
94.90	790.0
96.00	1771.0
97.00	837.0
98.00	30240.0
99.00	23200.0
100.00	1982.0
101.00	12516.0
102.00	681.0
103.00	4082.0
103.90	8259.0
105.00	6956.0
106.00	2282.0
107.00	96872.0
108.00	15575.0
109.00	2506.0
110.00	175872.0
111.00	27480.0
112.00	3381.0
112.90	1143.0
113.90	269.0
114.90	658.0
116.00	5264.0
117.00	90024.0
118.00	6843.0
119.00	1040.0
120.00	1427.0
121.00	515.0
122.00	6428.0
123.00	9871.0
124.00	4794.0
125.00	3976.0
127.00	338240.0
128.00	26304.0
129.00	128768.0
130.00	11783.0
131.00	2391.0

Instrument :
BNA_N
ClientSampleId :
DFTPP372

132.00	1408.0
132.90	1219.0
134.00	3668.0
135.00	11156.0
136.00	4043.0
137.00	6190.0
138.00	1264.0
138.90	895.0
140.00	1678.0
141.00	16680.0
142.00	5641.0
143.00	3893.0
143.90	1168.0
145.00	1087.0
146.00	2942.0
147.00	11212.0
148.00	22016.0
149.00	4848.0
150.00	1470.0
151.10	2050.0
151.80	1331.0
153.00	5258.0
154.00	4017.0
155.00	9334.0
156.00	14614.0
157.00	2834.0
157.90	2857.0
159.00	2549.0
160.00	5238.0
161.00	7860.0
162.00	2200.0
163.10	782.0
163.90	1047.0
165.00	6548.0
166.00	5050.0
167.00	37920.0
168.00	24136.0
169.00	3359.0
170.00	1309.0
170.90	1391.0
172.00	3398.0
173.00	4591.0
174.00	6924.0
175.00	13767.0
176.00	4027.0
177.00	6561.0
178.00	2262.0
178.90	25200.0
180.00	19368.0
181.00	8088.0
182.00	1515.0
182.90	798.0
184.00	2284.0
185.00	12892.0
186.00	94568.0
187.00	28432.0
188.00	3439.0
189.00	5839.0
190.00	1123.0
191.00	3427.0
192.00	9009.0
193.00	10241.0
194.00	2246.0
195.10	1206.0
196.00	22008.0
196.80	3754.0
198.00	731136.0
199.00	51360.0
200.00	3544.0
201.40	3424.0
203.00	4771.0
204.00	24408.0
205.00	43336.0
206.00	187200.0
207.10	29768.0
208.00	6733.0
209.00	2706.0
210.10	2971.0
211.00	7229.0
213.00	558.0
213.90	302.0
215.00	2328.0
216.00	3996.0
217.00	48776.0
218.00	6400.0
218.90	655.0
221.00	34424.0
223.00	10545.0
224.00	96328.0
225.00	27808.0
226.00	3275.0
227.00	37976.0
228.00	5811.0
229.00	8599.0
230.00	1125.0
231.00	3338.0

Instrument :
BNA_N
ClientSampleId :
DFTPP372

232.10	573.0
232.90	730.0
234.00	2778.0
235.00	2838.0
235.90	2312.0
237.00	4321.0
238.00	738.0
239.00	1528.0
239.90	1303.0
241.00	2218.0
242.00	5849.0
243.00	5672.0
244.00	83392.0
245.00	11745.0
246.00	14990.0
247.00	3395.0
247.90	910.0
249.00	3157.0
249.90	670.0
251.00	540.0
252.00	758.0
253.00	1715.0
254.10	3150.0
255.00	389824.0
256.00	58872.0
257.00	5127.0
258.00	23000.0
259.10	3613.0
260.00	707.0
260.90	752.0
263.00	281.0
264.00	588.0
265.00	9229.0
266.00	1445.0
266.90	718.0
267.90	182.0
269.90	524.0
270.90	862.0
272.00	973.0
273.00	12266.0
274.00	32472.0
275.00	184704.0
276.00	25136.0
277.00	14853.0
278.00	2892.0
279.00	681.0
281.10	3293.0
282.00	1069.0
283.00	2241.0
284.00	1193.0
285.00	2864.0
286.00	635.0
288.00	195.0
288.90	627.0
289.90	631.0
291.00	429.0
292.00	834.0
292.90	3371.0
294.00	1223.0
295.00	1342.0
296.00	49056.0
297.00	7094.0
297.90	370.0
298.90	173.0
301.00	761.0
302.00	1090.0
303.00	5209.0
304.00	1815.0
304.90	314.0
306.00	162.0
307.90	622.0
309.00	409.0
310.00	603.0
310.90	251.0
313.00	510.0
314.00	2144.0
315.00	5445.0
316.00	3226.0
317.00	538.0
320.00	284.0
321.00	1596.0
322.10	836.0
323.00	15640.0
324.10	3012.0
325.00	549.0
326.00	502.0
326.90	2921.0
328.00	1831.0
328.90	402.0
331.90	1045.0
333.00	1489.0
334.00	9929.0
335.00	2546.0
336.10	492.0
338.90	315.0
339.90	324.0

Instrument :
BNA_N
ClientSampleId :
DFTPP372

341.00	2598.0
342.10	748.0
343.00	522.0
345.90	3729.0
346.90	825.0
351.10	280.0
352.00	5149.0
353.00	3427.0
354.00	4493.0
355.00	2141.0
356.00	426.0
357.00	358.0
358.10	155.0
358.90	418.0
361.00	206.0
364.10	169.0
365.00	21960.0
366.00	3541.0
366.90	269.0
370.00	453.0
371.00	1365.0
372.00	7944.0
373.00	1885.0
374.00	237.0
376.80	178.0
382.90	2388.0
384.00	745.0
384.90	303.0
390.00	1140.0
391.00	836.0
391.90	625.0
399.00	157.0
401.10	577.0
402.00	3002.0
403.00	5043.0
404.00	1726.0
404.90	359.0
415.00	594.0
416.00	205.0
421.00	4505.0
422.00	4135.0
423.00	29464.0
424.00	9036.0
425.00	978.0
429.00	582.0
430.10	314.0
431.10	211.0
439.10	210.0
441.00	79832.0
442.00	467008.0
443.10	93664.0
444.00	9617.0
445.00	441.0
461.00	237.0
475.00	158.0
503.10	303.0

Instrument :
BNA_N
ClientSampleId :
DFTPP372

m/z	Abundance
35.00	329.0
37.00	1276.0
38.00	3181.0
39.00	17296.0
40.00	1415.0
41.00	751.0
42.00	217.0
43.00	516.0
44.00	1523.0
44.90	677.0
48.00	272.0
49.00	2085.0
50.00	62808.0
51.00	236992.0
52.00	11926.0
53.00	478.0
55.00	1652.0
56.00	7170.0
57.00	16279.0
57.90	784.0
59.00	415.0
59.90	215.0
61.00	3325.0
62.00	3759.0
63.00	10308.0
64.00	1566.0
65.00	5223.0
66.10	388.0
66.90	616.0
68.00	5198.0
69.00	286272.0
70.00	1583.0
70.80	335.0
72.00	236.0
73.00	5350.0
74.00	31384.0
75.00	47648.0
76.00	17016.0
77.00	350656.0
78.00	24936.0
79.00	22176.0
80.00	16176.0
81.00	23672.0
82.00	5822.0
83.00	5429.0
84.00	899.0
85.00	4165.0
86.00	7380.0
87.00	3461.0
88.00	1584.0
88.90	693.0
89.90	168.0
91.00	5522.0
92.00	5960.0
93.00	39912.0
94.00	2963.0
95.00	850.0
96.00	2174.0
97.00	883.0
98.00	30944.0
99.00	25168.0
100.00	1940.0
101.00	13347.0
102.00	945.0
103.00	4340.0
104.00	7916.0
105.00	8342.0
106.00	2732.0
107.00	105736.0
108.00	16135.0
109.00	3229.0
110.00	193536.0
111.00	29680.0
112.00	3775.0
112.90	1405.0
114.10	359.0
115.00	439.0
116.00	5941.0
117.00	99600.0
118.00	7500.0
119.00	918.0
120.00	1469.0
121.10	587.0
122.00	7389.0
123.00	11054.0
124.00	5136.0
125.00	4371.0
126.10	1831.0
127.00	375744.0
128.00	28480.0
129.00	141504.0
130.00	12374.0

Instrument :
BNA_N
ClientSampleId :
DFTPP372

131.00	2709.0
132.00	1309.0
132.90	882.0
134.00	4604.0
135.00	11782.0
136.00	5079.0
137.00	6973.0
138.00	1433.0
139.00	914.0
140.00	1762.0
141.00	19456.0
142.00	6805.0
142.90	4590.0
144.00	1325.0
145.00	1125.0
146.00	3700.0
147.00	11173.0
148.00	24952.0
149.00	5503.0
150.00	1428.0
151.10	2720.0
151.90	1398.0
153.00	6009.0
154.00	5383.0
155.00	10863.0
156.00	15571.0
157.00	3295.0
158.00	3493.0
159.00	2701.0
160.00	5929.0
161.00	9291.0
162.00	2402.0
163.00	940.0
164.00	1141.0
165.00	7449.0
166.00	6179.0
167.00	42992.0
168.00	25744.0
169.00	3503.0
170.00	1364.0
170.90	1584.0
172.00	3575.0
173.00	4802.0
174.00	8900.0
175.00	14704.0
176.10	4565.0
177.00	7425.0
178.00	2644.0
179.00	29312.0
180.00	21376.0
181.00	10081.0
182.00	1692.0
182.90	809.0
184.00	2235.0
185.00	14383.0
186.00	113480.0
187.00	32112.0
188.00	3726.0
189.00	7346.0
190.00	1248.0
191.00	3870.0
192.00	10677.0
193.00	10829.0
194.00	2458.0
194.90	1506.0
196.00	25080.0
197.10	5115.0
198.00	887424.0
199.00	58016.0
200.00	4636.0
201.40	3937.0
203.00	5266.0
204.00	30128.0
205.00	50896.0
206.10	218368.0
207.00	32152.0
208.00	7781.0
209.00	2750.0
210.20	3877.0
211.00	9025.0
211.80	1936.0
213.00	547.0
213.90	317.0
215.00	2652.0
216.00	5042.0
216.90	56768.0
218.00	7678.0
219.10	769.0
221.00	41992.0
221.80	9647.0
223.00	13074.0
224.10	126176.0
225.00	33872.0
226.00	3599.0
227.00	50520.0
228.00	7261.0

Instrument :
BNA_N
ClientSampleId :
DFTPP372

229.00	11405.0
230.00	1528.0
231.00	4082.0
232.00	790.0
233.00	874.0
234.00	3374.0
235.00	3509.0
236.00	2092.0
237.00	4924.0
238.00	697.0
239.00	1844.0
240.00	1616.0
241.00	2960.0
242.00	6743.0
243.00	7040.0
244.00	107984.0
245.10	13855.0
246.00	19648.0
247.00	3688.0
248.00	932.0
249.00	3579.0
250.00	676.0
250.90	1015.0
252.00	965.0
253.00	1917.0
254.00	3520.0
255.00	493312.0
256.00	76688.0
257.00	6653.0
258.00	27552.0
259.10	4263.0
259.90	1020.0
260.90	860.0
262.00	156.0
262.90	345.0
263.90	683.0
265.00	11379.0
265.90	1810.0
266.90	677.0
268.00	258.0
269.00	201.0
269.90	505.0
271.00	1187.0
271.90	1337.0
273.00	15413.0
274.00	43040.0
275.00	248064.0
276.00	32352.0
277.00	19152.0
278.00	3859.0
279.00	654.0
279.90	189.0
281.00	2203.0
282.00	873.0
283.00	2647.0
284.00	1430.0
285.00	3502.0
285.90	665.0
288.00	205.0
288.90	866.0
290.00	776.0
291.00	460.0
291.90	873.0
293.00	4526.0
293.90	1333.0
295.10	1361.0
296.00	63928.0
297.00	9495.0
297.90	716.0
299.00	196.0
301.00	872.0
302.00	959.0
303.00	7903.0
304.00	2278.0
304.90	283.0
308.00	887.0
308.90	533.0
310.00	721.0
310.80	272.0
311.90	250.0
313.00	610.0
314.00	2816.0
315.00	6879.0
316.00	4197.0
317.00	782.0
318.90	161.0
319.90	205.0
320.90	2075.0
322.00	1170.0
323.00	21216.0
324.10	3734.0
325.00	736.0
325.90	636.0
327.00	4701.0
327.90	2217.0
329.00	510.0

Instrument :
BNA_N
ClientSampleId :
DFTPP372

332.00	1609.0
333.00	2256.0
334.00	14778.0
335.00	3540.0
336.00	398.0
338.90	449.0
339.90	461.0
341.00	3047.0
342.00	964.0
343.10	220.0
346.00	5919.0
346.90	997.0
347.90	168.0
350.90	450.0
352.00	7181.0
352.90	5180.0
354.00	7022.0
355.10	2092.0
356.00	331.0
357.00	233.0
357.90	224.0
359.00	627.0
359.90	171.0
364.10	359.0
365.00	29408.0
366.00	5034.0
367.00	375.0
370.00	563.0
370.90	1880.0
372.00	12315.0
373.00	2753.0
374.00	327.0
377.00	368.0
382.90	3371.0
384.00	974.0
384.90	395.0
390.00	1569.0
390.90	1113.0
392.00	911.0
393.10	175.0
397.00	247.0
400.90	754.0
402.00	4972.0
403.00	7709.0
404.10	2815.0
405.00	598.0
409.90	294.0
415.00	487.0
415.80	160.0
421.00	6289.0
422.00	6768.0
423.00	42552.0
424.00	10114.0
425.00	1549.0
428.90	218.0
429.10	215.0
439.10	197.0
441.10	116440.0
442.00	796608.0
443.10	135488.0
444.00	13513.0
445.00	714.0
461.00	173.0
503.20	161.0

Instrument :
BNA_N
ClientSampleId :
DFTPP372

m/z	Abundance
35.90	178.0
37.00	696.0
38.00	1967.0
39.00	11612.0
40.00	1346.0
41.10	877.0
42.00	200.0
43.00	556.0
44.00	1459.0
45.00	493.0
46.90	188.0
47.90	191.0
49.00	1384.0
50.00	41416.0
51.00	152896.0
52.00	7757.0
53.00	516.0
55.00	1314.0
56.00	4678.0
57.00	11506.0
57.90	620.0
59.00	263.0
60.00	191.0
61.00	2393.0
62.00	2687.0
63.00	7092.0
64.00	1049.0
65.00	3881.0
65.90	471.0
67.00	524.0
68.00	3258.0
69.00	190720.0
70.00	1151.0
71.00	302.0
73.00	3377.0
74.00	20680.0
75.00	31488.0
76.00	10776.0
77.00	231680.0
78.00	16190.0
79.00	14817.0
80.00	10828.0
81.00	16126.0
82.00	3973.0
83.00	3838.0
83.90	655.0
85.00	3135.0
86.00	4930.0
87.00	2247.0
87.90	1087.0
89.10	498.0
91.00	3895.0
92.00	4207.0
93.00	28200.0
93.90	2020.0
95.00	606.0
96.10	1584.0
97.10	540.0
98.00	22016.0
99.00	17400.0
100.00	1496.0
101.00	9289.0
101.90	515.0
103.00	2989.0
104.00	5394.0
104.90	5867.0
106.00	1861.0
107.00	71448.0
108.00	11015.0
109.00	2184.0
110.00	132032.0
111.00	20000.0
111.90	2505.0
113.00	814.0
114.10	294.0
115.00	492.0
116.00	4085.0
117.00	66840.0
118.00	4869.0
119.00	801.0
120.00	1208.0
121.00	487.0
122.00	5061.0
123.00	7484.0
123.90	3444.0
124.90	3205.0
126.10	1116.0
127.00	266304.0
128.00	18752.0
129.00	100680.0
130.00	8464.0
131.00	1845.0

Instrument :
BNA_N
ClientSampleId :
DFTPP372

132.00	918.0
132.90	834.0
134.00	3093.0
135.00	8627.0
136.00	3408.0
137.00	5231.0
137.90	1107.0
138.90	703.0
140.00	1415.0
140.90	13723.0
142.00	4364.0
143.00	3169.0
143.90	907.0
145.00	1085.0
146.00	2585.0
147.00	7368.0
148.00	17512.0
149.00	4037.0
150.00	1138.0
151.10	1785.0
151.90	1170.0
153.00	4016.0
154.00	3325.0
155.00	7580.0
156.00	11400.0
157.00	2107.0
157.90	2552.0
159.00	1921.0
160.00	4340.0
161.00	6697.0
162.00	1969.0
162.90	706.0
164.00	682.0
165.00	5257.0
166.00	3914.0
167.00	31976.0
168.00	18680.0
169.00	2563.0
170.00	1066.0
170.90	1423.0
172.00	2809.0
173.00	3450.0
174.00	6591.0
175.00	11607.0
176.00	3015.0
177.00	5413.0
178.00	1728.0
178.90	21232.0
180.00	15582.0
181.00	6889.0
182.00	1177.0
182.80	530.0
183.10	530.0
184.00	1726.0
185.00	10593.0
186.00	82288.0
187.00	22104.0
188.00	2634.0
189.00	5352.0
189.90	1010.0
191.00	2997.0
192.00	7670.0
193.00	8418.0
194.10	2012.0
195.00	864.0
196.00	17440.0
197.00	3010.0
198.00	640000.0
199.00	40120.0
200.00	2959.0
201.50	3297.0
203.00	4096.0
204.00	21984.0
205.00	38416.0
206.10	162944.0
207.00	25168.0
208.00	6268.0
209.00	2048.0
210.20	2743.0
211.00	6221.0
213.00	562.0
214.00	155.0
215.00	2148.0
216.00	3391.0
217.00	43984.0
218.00	5764.0
219.00	546.0
221.00	29072.0
221.80	6019.0
223.00	8967.0
224.00	90376.0
225.00	22616.0
226.00	2570.0
227.00	37640.0
228.00	4918.0
229.00	7949.0

Instrument :
BNA_N
ClientSampleId :
DFTPP372

229.90	1023.0
231.00	3299.0
232.00	469.0
233.00	844.0
233.90	2485.0
235.00	2672.0
236.00	1831.0
237.10	3727.0
238.00	695.0
238.90	1413.0
239.90	1073.0
241.00	2302.0
242.00	5059.0
243.10	5537.0
244.00	79256.0
245.00	9673.0
246.00	14059.0
247.00	2772.0
247.90	619.0
249.00	2964.0
250.00	543.0
250.90	610.0
251.90	784.0
253.00	1526.0
255.00	382656.0
256.00	53616.0
257.00	3992.0
258.00	21464.0
259.00	3490.0
260.00	417.0
261.00	670.0
263.10	180.0
264.00	693.0
265.00	9335.0
265.90	1226.0
266.90	397.0
267.90	279.0
269.00	161.0
270.00	428.0
270.90	964.0
272.00	1036.0
273.00	11936.0
274.00	33600.0
275.00	191296.0
276.00	23144.0
277.00	15316.0
277.90	2576.0
279.00	500.0
281.00	1365.0
282.10	700.0
283.00	1685.0
284.00	1221.0
285.10	2853.0
286.00	498.0
287.90	158.0
289.00	581.0
290.10	616.0
290.90	418.0
292.00	660.0
293.00	3688.0
294.00	900.0
296.00	48048.0
297.00	6910.0
298.10	458.0
298.90	165.0
301.00	721.0
302.00	857.0
303.00	5967.0
304.00	1991.0
304.90	262.0
308.00	696.0
309.00	535.0
310.00	656.0
311.00	231.0
311.90	179.0
312.90	538.0
314.00	2270.0
315.00	5587.0
316.00	3113.0
317.00	555.0
320.00	229.0
321.00	1655.0
322.00	821.0
323.00	18456.0
324.00	2997.0
325.10	342.0
326.00	379.0
327.00	3414.0
328.00	1508.0
329.10	240.0
331.90	1309.0
333.00	1770.0
334.00	11807.0
335.00	2986.0
335.90	318.0
338.80	207.0

Instrument :
BNA_N
ClientSampleId :
DFTPP372

339.10	202.0
339.90	220.0
341.00	2325.0
342.00	569.0
345.90	4424.0
346.90	770.0
351.00	270.0
352.00	5316.0
353.00	4048.0
354.00	5508.0
355.00	1436.0
356.00	201.0
359.00	346.0
365.00	26688.0
366.00	4193.0
366.90	344.0
369.90	507.0
371.00	1716.0
372.00	10014.0
373.00	2793.0
374.00	311.0
376.80	238.0
383.10	2769.0
384.00	914.0
384.90	309.0
390.10	1616.0
391.00	894.0
392.00	730.0
396.90	207.0
400.90	739.0
401.90	4256.0
403.00	6774.0
404.00	2503.0
405.10	390.0
410.00	205.0
415.10	363.0
420.20	178.0
421.00	5731.0
422.00	5408.0
423.10	40576.0
424.00	9462.0
425.00	1055.0
429.00	163.0
438.40	155.0
438.90	208.0
440.00	610.0
441.00	100312.0
442.00	673984.0
443.00	130216.0
444.00	10929.0
444.90	570.0

Instrument :
BNA_N
ClientSampleId :
DFTPP372