

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090925\
 Data File : BN037674.D
 Acq On : 09 Sep 2025 20:45
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN090925

Quant Time: Sep 10 04:11:29 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN090925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 10 02:02:04 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.847	152	5543	0.400	ng	0.00
7) Naphthalene-d8	10.648	136	15284	0.400	ng	0.00
13) Acenaphthene-d10	14.495	164	7827	0.400	ng	0.00
19) Phenanthrene-d10	17.235	188	13881	0.400	ng	#-0.01
29) Chrysene-d12	21.411	240	11300	0.400	ng	# 0.00
35) Perylene-d12	23.744	264	11536	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.406	112	5520	0.389	ng	0.00
5) Phenol-d6	6.988	99	7108	0.399	ng	0.00
8) Nitrobenzene-d5	8.993	82	4325	0.459	ng	0.00
11) 2-Methylnaphthalene-d10	12.243	152	9174	0.427	ng	0.00
14) 2,4,6-Tribromophenol	15.982	330	830	0.381	ng	0.00
15) 2-Fluorobiphenyl	13.126	172	15027	0.488	ng	0.00
27) Fluoranthene-d10	19.271	212	14894	0.416	ng	0.00
31) Terphenyl-d14	19.870	244	11595	0.487	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.304	88	2539	0.422	ng	98
3) n-Nitrosodimethylamine	3.608	42	3050	0.403	ng	# 96
6) bis(2-Chloroethyl)ether	7.262	93	6313	0.419	ng	99
9) Naphthalene	10.701	128	16929	0.409	ng	100
10) Hexachlorobutadiene	11.000	225	3105	0.424	ng	# 98
12) 2-Methylnaphthalene	12.319	142	10166	0.379	ng	99
16) Acenaphthylene	14.217	152	16164	0.442	ng	100
17) Acenaphthene	14.559	154	9742	0.415	ng	99
18) Fluorene	15.535	166	11826	0.412	ng	99
20) 4,6-Dinitro-2-methylph...	15.609	198	314	0.406	ng	# 72
21) 4-Bromophenyl-phenylether	16.441	248	3811	0.419	ng	94
22) Hexachlorobenzene	16.553	284	4243	0.442	ng	98
23) Atrazine	16.702	200	2371	0.417	ng	98
24) Pentachlorophenol	16.888	266	1057	0.328	ng	97
25) Phenanthrene	17.285	178	17882	0.421	ng	100
26) Anthracene	17.372	178	16641	0.423	ng	99
28) Fluoranthene	19.299	202	19002	0.399	ng	100
30) Pyrene	19.657	202	18987	0.421	ng	99
32) Benzo(a)anthracene	21.393	228	17190	0.426	ng	99
33) Chrysene	21.447	228	18012	0.440	ng	99
34) Bis(2-ethylhexyl)phtha...	21.349	149	4405	0.393	ng	# 99
36) Indeno(1,2,3-cd)pyrene	26.139	276	19883	0.455	ng	99
37) Benzo(b)fluoranthene	23.043	252	18236	0.429	ng	99
38) Benzo(k)fluoranthene	23.090	252	18290	0.437	ng	100
39) Benzo(a)pyrene	23.642	252	16146	0.456	ng	100
40) Dibenzo(a,h)anthracene	26.157	278	15807	0.481	ng	97
41) Benzo(g,h,i)perylene	26.864	276	17019	0.429	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090925\
Data File : BN037674.D
Acq On : 09 Sep 2025 20:45
Operator : RC/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN090925


Quant Time: Sep 10 04:11:29 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN090925.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 10 02:02:04 2025
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.847 min Scan# 660
 Delta R.T. 0.000 min
 Lab File: BN037674.D
 Acq: 09 Sep 2025 20:45


Instrument :
 BNA_N
 ClientSampleId :
 ICVBN090925





Tgt Ion:152 Resp: 5543
 Ion Ratio Lower Upper
 152 100
 150 149.9 123.2 184.8
 115 59.4 49.0 73.4




#2
 1,4-Dioxane
 Concen: 0.422 ng
 RT: 3.304 min Scan# 31
 Delta R.T. 0.000 min
 Lab File: BN037674.D
 Acq: 09 Sep 2025 20:45



Tgt Ion: 88 Resp: 2539
 Ion Ratio Lower Upper
 88 100
 43 36.2 26.8 40.2
 58 77.4 62.6 93.8









#3
n-Nitrosodimethylamine
Concen: 0.403 ng
RT: 3.608 min Scan# 73
Delta R.T. 0.000 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45




Instrument :
BNA_N
ClientSampleId :
ICVBN090925


Tgt Ion: 42 Resp: 3050
Ion Ratio Lower Upper
42 100
74 121.6 94.9 142.3
44 7.2 10.6 16.0#



#4
2-Fluorophenol
Concen: 0.389 ng
RT: 5.406 min Scan# 322
Delta R.T. 0.000 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45


Tgt Ion: 112 Resp: 5520
Ion Ratio Lower Upper
112 100
64 57.0 45.4 68.2
63 29.9 23.7 35.5








#5
Phenol-d6
Concen: 0.399 ng
RT: 6.988 min Scan# 541
Delta R.T. 0.000 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45

Instrument :
BNA_N
ClientSampleId :
ICVBN090925



Tgt Ion: 99 Resp: 7108




Ion	Ratio	Lower	Upper
99	100		
42	21.2	17.7	26.5
71	32.4	26.7	40.1




#6
bis(2-Chloroethyl)ether
Concen: 0.419 ng
RT: 7.262 min Scan# 579
Delta R.T. 0.000 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45

Tgt Ion: 93 Resp: 6313

Ion	Ratio	Lower	Upper
93	100		
63	76.4	62.1	93.1
95	32.9	26.0	39.0









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.648 min Scan# 943
Delta R.T. 0.000 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45




Instrument :
BNA_N
ClientSampleId :
ICVBN090925


Tgt Ion:	136	Resp:	15284
Ion	Ratio	Lower	Upper
136	100		
137	11.5	9.3	13.9
54	8.9	8.1	12.1
68	6.9	6.2	9.2



#8
Nitrobenzene-d5
Concen: 0.459 ng
RT: 8.993 min Scan# 788
Delta R.T. 0.000 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45

Tgt Ion:	82	Resp:	4325
Ion	Ratio	Lower	Upper
82	100		
128	44.9	34.3	51.5
54	55.5	45.5	68.3





#9

Naphthalene

Concen: 0.409 ng

RT: 10.701 min Scan# 948

Delta R.T. 0.000 min

Lab File: BN037674.D

Acq: 09 Sep 2025 20:45

Instrument :

BNA_N

ClientSampleId :

ICVBN090925





Tgt Ion:128 Resp: 16929

Ion Ratio Lower Upper

128 100

129 11.9 9.3 13.9

127 13.4 10.7 16.1



#10

Hexachlorobutadiene

Concen: 0.424 ng

RT: 11.000 min Scan# 976

Delta R.T. 0.000 min

Lab File: BN037674.D

Acq: 09 Sep 2025 20:45




Tgt Ion:225 Resp: 3105


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 63.2 51.7 77.5









#11
2-Methylnaphthalene-d10
Concen: 0.427 ng
RT: 12.243 min Scan# 1168
Delta R.T. 0.000 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45




Instrument :
BNA_N
ClientSampleId :
ICVBN090925


Tgt Ion:152 Resp: 9174
Ion Ratio Lower Upper
152 100
151 20.7 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.379 ng
RT: 12.319 min Scan# 1183
Delta R.T. 0.000 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45

Tgt Ion:142 Resp: 10166
Ion Ratio Lower Upper
142 100
141 88.4 71.5 107.3
115 34.6 28.0 42.0









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.495 min Scan# 1458
Delta R.T. 0.000 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45




Instrument :
BNA_N
ClientSampleId :
ICVBN090925


Tgt Ion:164 Resp: 7827
Ion Ratio Lower Upper
164 100
162 103.3 82.6 123.8
160 50.3 39.7 59.5



#14
2,4,6-Tribromophenol
Concen: 0.381 ng
RT: 15.982 min Scan# 1591
Delta R.T. 0.000 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45

Tgt Ion:330 Resp: 830
Ion Ratio Lower Upper
330 100
332 94.2 78.8 118.2
141 41.6 37.8 56.6









#15
2-Fluorobiphenyl
Concen: 0.488 ng
RT: 13.126 min Scan# 1330
Delta R.T. 0.000 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45




Instrument :
BNA_N
ClientSampleId :
ICVBN090925


Tgt Ion:172 Resp: 15027
Ion Ratio Lower Upper
172 100
171 34.6 27.8 41.8
170 22.6 18.4 27.6



#16
Acenaphthylene
Concen: 0.442 ng
RT: 14.217 min Scan# 1432
Delta R.T. 0.000 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45

Tgt Ion:152 Resp: 16164
Ion Ratio Lower Upper
152 100
151 20.1 16.0 24.0
153 13.0 10.4 15.6









#17
Acenaphthene
Concen: 0.415 ng
RT: 14.559 min Scan# 1464
Delta R.T. 0.000 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45




Instrument :
BNA_N
ClientSampleId :
ICVBN090925


Tgt Ion:154 Resp: 9742
Ion Ratio Lower Upper
154 100
153 112.4 90.7 136.1
152 58.7 46.4 69.6



#18
Fluorene
Concen: 0.412 ng
RT: 15.535 min Scan# 1555
Delta R.T. -0.012 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45

Tgt Ion:166 Resp: 11826
Ion Ratio Lower Upper
166 100
165 97.5 77.4 116.0
167 13.3 10.6 15.8









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.235 min Scan# 1693
Delta R.T. -0.012 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45




Instrument :
BNA_N
ClientSampleId :
ICVBN090925


Tgt Ion:188 Resp: 13881
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 13.7 6.3 9.5#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.406 ng
RT: 15.609 min Scan# 1561
Delta R.T. 0.000 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45

Tgt Ion:198 Resp: 314
Ion Ratio Lower Upper
198 100
51 144.5 144.8 217.2#
105 101.7 110.7 166.1#









#21
4-Bromophenyl-phenylether
Concen: 0.419 ng
RT: 16.441 min Scan# 1628
Delta R.T. 0.000 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45




Instrument :
BNA_N
ClientSampleId :
ICVBN090925


Tgt Ion	Ratio	Lower	Upper
248	100		
250	102.9	78.7	118.1
141	45.0	40.2	60.4



#22
Hexachlorobenzene
Concen: 0.442 ng
RT: 16.553 min Scan# 1637
Delta R.T. 0.000 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45

Tgt Ion	Ratio	Lower	Upper
284	100		
142	37.9	29.4	44.2
249	31.4	24.4	36.6









#23
 Atrazine
 Concen: 0.417 ng
 RT: 16.702 min Scan# 1649
 Delta R.T. 0.000 min
 Lab File: BN037674.D
 Acq: 09 Sep 2025 20:45




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN090925


Tgt Ion:200 Resp: 2371
 Ion Ratio Lower Upper
 200 100
 173 28.4 24.7 37.1
 215 47.1 37.5 56.3



#24
 Pentachlorophenol
 Concen: 0.328 ng
 RT: 16.888 min Scan# 1664
 Delta R.T. -0.012 min
 Lab File: BN037674.D
 Acq: 09 Sep 2025 20:45

Tgt Ion:266 Resp: 1057
 Ion Ratio Lower Upper
 266 100
 264 61.4 49.7 74.5
 268 64.7 48.7 73.1









#25
Phenanthrene
Concen: 0.421 ng
RT: 17.285 min Scan# 1696
Delta R.T. 0.000 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45




Instrument :
BNA_N
ClientSampleId :
ICVBN090925


Tgt Ion:178 Resp: 17882
Ion Ratio Lower Upper
178 100
176 19.3 15.3 22.9
179 15.5 12.5 18.7



#26
Anthracene
Concen: 0.423 ng
RT: 17.372 min Scan# 1703
Delta R.T. 0.000 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45

Tgt Ion:178 Resp: 16641
Ion Ratio Lower Upper
178 100
176 18.8 14.8 22.2
179 15.6 12.0 18.0









#27
Fluoranthene-d10
Concen: 0.416 ng
RT: 19.271 min Scan# 1958
Delta R.T. 0.000 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45




Instrument :
BNA_N
ClientSampleId :
ICVBN090925


Tgt Ion:212 Resp: 14894
Ion Ratio Lower Upper
212 100
106 15.1 11.8 17.6
104 8.6 6.6 10.0



#28
Fluoranthene
Concen: 0.399 ng
RT: 19.299 min Scan# 1964
Delta R.T. 0.000 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45

Tgt Ion:202 Resp: 19002
Ion Ratio Lower Upper
202 100
101 11.3 9.0 13.6
203 17.1 13.8 20.8





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.411 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN037674.D

Acq: 09 Sep 2025 20:45

Instrument :

BNA_N

ClientSampleId :

ICVBN090925





Tgt Ion:240 Resp: 11300

Ion Ratio Lower Upper

240 100

120 15.9 8.5 12.7#

236 28.3 22.5 33.7



#30

Pyrene

Concen: 0.421 ng

RT: 19.657 min Scan# 2041

Delta R.T. -0.005 min

Lab File: BN037674.D

Acq: 09 Sep 2025 20:45




Tgt Ion:202 Resp: 18987


Ion Ratio Lower Upper

202 100

200 20.8 16.8 25.2

203 17.5 14.3 21.5





#31

Terphenyl-d14

Concen: 0.487 ng

RT: 19.870 min Scan# 2087

Delta R.T. 0.000 min

Lab File: BN037674.D

Acq: 09 Sep 2025 20:45

Instrument :

BNA_N

ClientSampleId :

ICVBN090925





Tgt Ion:244 Resp: 11595

Ion Ratio Lower Upper

244 100

212 8.8 7.0 10.4

122 15.4 13.4 20.2



#32

Benzo(a)anthracene

Concen: 0.426 ng

RT: 21.393 min Scan# 2271

Delta R.T. -0.009 min

Lab File: BN037674.D

Acq: 09 Sep 2025 20:45




Tgt Ion:228 Resp: 17190


Ion Ratio Lower Upper

228 100

226 27.7 21.8 32.6

229 19.6 15.8 23.6





#33

Chrysene

Concen: 0.440 ng

RT: 21.447 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN037674.D

Acq: 09 Sep 2025 20:45

Instrument :

BNA_N

ClientSampleId :

ICVBN090925





Tgt Ion:228 Resp: 18012

Ion Ratio Lower Upper

228 100

226 30.5 23.5 35.3

229 19.9 16.1 24.1



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.393 ng

RT: 21.349 min Scan# 2266

Delta R.T. -0.009 min

Lab File: BN037674.D

Acq: 09 Sep 2025 20:45




Tgt Ion:149 Resp: 4405


Ion Ratio Lower Upper

149 100

167 27.0 21.4 32.0

279 4.3 2.6 3.8#









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.744 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: BN037674.D
 Acq: 09 Sep 2025 20:45




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN090925


Tgt Ion:264 Resp: 11536
 Ion Ratio Lower Upper
 264 100
 260 27.0 21.2 31.8
 265 38.7 28.2 42.4



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.455 ng
 RT: 26.139 min Scan# 3661
 Delta R.T. -0.003 min
 Lab File: BN037674.D
 Acq: 09 Sep 2025 20:45


Tgt Ion:276 Resp: 19883
 Ion Ratio Lower Upper
 276 100
 138 26.8 21.3 31.9
 277 25.4 20.1 30.1








#37
Benzo(b)fluoranthene
Concen: 0.429 ng
RT: 23.043 min Scan# 2602
Delta R.T. -0.003 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45

Instrument :
BNA_N
ClientSampleId :
ICVBN090925







Tgt Ion:252 Resp: 18236
Ion Ratio Lower Upper
252 100
253 25.2 19.7 29.5
125 13.0 10.2 15.2



#38
Benzo(k)fluoranthene
Concen: 0.437 ng
RT: 23.090 min Scan# 2618
Delta R.T. -0.006 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45

Tgt Ion:252 Resp: 18290
Ion Ratio Lower Upper
252 100
253 24.8 19.9 29.9
125 13.9 11.4 17.0









#39
Benzo(a)pyrene
Concen: 0.456 ng
RT: 23.642 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45




Instrument :
BNA_N
ClientSampleId :
ICVBN090925


Tgt Ion:	252	Resp:	16146
Ion Ratio	Lower	Upper	
252	100		
253	26.1	20.7	31.1
125	14.7	11.9	17.9



#40
Dibenzo(a,h)anthracene
Concen: 0.481 ng
RT: 26.157 min Scan# 3667
Delta R.T. -0.009 min
Lab File: BN037674.D
Acq: 09 Sep 2025 20:45

Tgt Ion:	278	Resp:	15807
Ion Ratio	Lower	Upper	
278	100		
139	21.5	19.1	28.7
279	26.4	20.9	31.3





#41
 Benzo(g,h,i)perylene
 Concen: 0.429 ng
 RT: 26.864 min Scan# 3911
 Delta R.T. -0.006 min
 Lab File: BN037674.D
 Acq: 09 Sep 2025 20:45

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN090925

Tgt Ion:	276	Resp:	17019
Ion	Ratio	Lower	Upper
276	100		
277	24.9	19.9	29.9
138	25.2	20.5	30.7

