


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN091018\
Data File : BN002614.D
Acq On : 10 Sep 2018 09:57
Operator : JU/SJ
Sample : DFTPP14
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP14

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN081318MA.M
Title : SVOA CALIBRATION
Last Update : Mon Sep 10 17:47:48 2018



AutoFind: Scans 2465, 2466, 2467; Background Corrected with Scan 2459

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.2	112816	PASS
68	69	0.00	2	1.8	2121	PASS
69	198	0.00	100	38.2	115713	PASS
70	69	0.00	2	0.7	781	PASS
127	198	10	80	45.8	138672	PASS
197	198	0.00	2	0.3	935	PASS
198	198	100	100	100.0	303082	PASS
199	198	5	9	6.6	19952	PASS
275	198	10	60	26.1	79120	PASS
365	198	1	100	3.1	9286	PASS
441	198	0.01	100	13.8	41936	PASS
442	198	50	100	89.2	270464	PASS
443	442	15	24	19.4	52474	PASS

m/z	Abundance
37.10	339.0
38.00	1175.0
39.00	7472.0
40.00	813.0
41.10	953.0
43.10	921.0
44.00	7241.0
44.90	361.0
49.00	696.0
50.00	24928.0
51.00	97624.0
52.00	5602.0
53.00	357.0
55.00	1279.0
56.00	3382.0
57.00	7681.0
58.00	397.0
61.00	1344.0
62.00	1404.0
63.00	3935.0
64.00	615.0
65.00	1916.0
66.00	314.0
67.10	450.0
68.00	1503.0
69.00	96528.0
70.10	580.0
71.00	350.0
73.00	1532.0
74.00	8679.0
75.00	14906.0
76.00	4506.0
77.00	109072.0
78.00	7366.0
79.00	7027.0
80.00	5181.0
81.00	7890.0
82.00	1872.0
83.00	2363.0
85.00	1490.0
86.00	1892.0
86.90	1175.0
87.80	335.0
91.00	1849.0
92.00	2051.0
93.00	10944.0
93.90	1014.0
95.00	576.0
95.90	864.0
97.10	498.0
98.00	8470.0
99.00	6553.0
99.90	433.0
101.00	4211.0
101.80	440.0
103.00	1526.0
104.00	2244.0
105.00	2475.0
106.00	845.0
107.00	30904.0
108.00	4664.0
110.00	57856.0
111.00	8912.0
112.00	1120.0
113.00	396.0
114.90	401.0
116.00	1993.0
117.00	20936.0
118.00	1315.0
119.00	575.0
121.00	413.0
122.00	2346.0
123.00	3155.0
123.90	1356.0
125.00	1151.0
127.00	111760.0
128.00	8496.0
129.00	40608.0
129.90	3843.0
131.10	654.0
131.90	531.0
132.60	359.0
133.10	380.0
134.00	1221.0
135.00	3879.0
136.00	1548.0
136.90	2002.0
138.00	458.0
140.00	704.0
140.90	5409.0
141.90	1570.0

Instrument :
BNA_N
ClientSampleId :
DFTPP14

143.00	1356.0
144.00	439.0
145.00	481.0
145.90	1067.0
147.00	3140.0
148.00	6065.0
149.00	1416.0
149.90	408.0
151.20	849.0
151.80	391.0
153.00	1533.0
153.90	1506.0
155.00	3083.0
156.00	4108.0
157.00	996.0
157.90	935.0
158.90	848.0
160.00	1520.0
161.00	2530.0
161.90	821.0
163.00	546.0
164.10	431.0
165.00	2381.0
165.90	1683.0
167.00	9929.0
167.90	4563.0
169.00	977.0
170.10	457.0
170.90	753.0
172.00	988.0
173.00	1127.0
173.90	2623.0
175.00	4489.0
176.00	1191.0
176.90	2642.0
178.00	804.0
178.90	7289.0
180.00	5506.0
181.00	2427.0
181.90	524.0
184.00	726.0
185.00	3545.0
186.00	29360.0
187.00	9042.0
188.00	1220.0
189.00	2102.0
190.00	532.0
191.00	1011.0
192.00	2647.0
193.00	2551.0
194.00	761.0
194.90	384.0
196.00	6252.0
196.90	1179.0
197.90	223744.0
199.00	14696.0
200.00	1391.0
201.30	1097.0
202.90	1423.0
204.00	7180.0
205.00	12949.0
206.00	56272.0
207.00	9383.0
207.90	1974.0
208.90	779.0
210.80	2301.0
211.80	437.0
215.00	505.0
216.00	1298.0
216.90	13614.0
218.00	1580.0
219.90	336.0
221.00	9480.0
223.00	3281.0
224.00	30696.0
225.00	7909.0
226.00	987.0
227.00	12460.0
227.90	1814.0
228.90	2737.0
229.90	348.0
231.00	1453.0
231.90	309.0
233.90	793.0
235.00	905.0
235.90	683.0
236.90	900.0
238.90	556.0
239.90	343.0
240.90	674.0
241.90	1480.0
243.00	1648.0
244.00	21328.0
245.00	2817.0
246.00	4410.0

Instrument :
BNA_N
ClientSampleId :
DFTPP14

247.00	1127.0
249.00	965.0
253.00	516.0
255.00	117944.0
256.00	17512.0
257.00	1132.0
257.90	6319.0
258.90	1079.0
265.00	2479.0
266.00	465.0
271.90	317.0
273.00	3835.0
274.00	7784.0
275.00	51728.0
276.00	7134.0
277.00	4000.0
277.90	908.0
282.90	547.0
283.90	304.0
285.00	812.0
292.90	907.0
293.90	359.0
296.00	12652.0
297.00	2264.0
301.00	386.0
301.80	336.0
303.00	1656.0
303.90	485.0
313.90	532.0
314.90	1451.0
316.00	1151.0
320.80	394.0
322.00	352.0
323.00	4862.0
324.00	954.0
326.90	985.0
327.70	334.0
328.00	400.0
333.00	442.0
334.00	2798.0
335.10	710.0
341.00	661.0
346.00	832.0
352.00	1235.0
353.00	876.0
353.90	1320.0
355.00	461.0
365.00	5906.0
365.90	973.0
372.00	2198.0
372.90	532.0
382.90	826.0
390.70	311.0
402.00	725.0
403.10	1081.0
403.90	479.0
420.90	953.0
422.00	893.0
423.00	7805.0
424.10	1824.0
425.00	467.0
441.00	22256.0
442.00	145728.0
443.00	29592.0
444.00	3187.0

Instrument :
BNA_N
ClientSampleId :
DFTPP14

m/z	Abundance
37.00	387.0
38.00	1683.0
39.00	10228.0
40.00	913.0
41.00	953.0
43.10	657.0
44.00	7367.0
45.00	431.0
49.00	1033.0
50.00	35192.0
51.00	134272.0
52.00	6585.0
53.10	430.0
55.00	1055.0
56.00	4206.0
57.00	10464.0
57.90	660.0
61.00	2014.0
62.00	1983.0
63.00	5276.0
64.00	917.0
65.10	2652.0
67.00	399.0
68.10	2483.0
69.00	139520.0
70.00	945.0
71.20	320.0
73.00	1824.0
74.00	13359.0
75.00	20384.0
76.00	7038.0
77.00	152384.0
78.00	10541.0
79.00	10459.0
80.00	6939.0
81.00	11005.0
82.00	3050.0
83.00	2719.0
85.00	2140.0
86.00	2857.0
87.00	1361.0
88.00	641.0
89.10	316.0
91.00	2814.0
92.00	2531.0
93.00	16408.0
94.00	1295.0
95.10	698.0
96.00	849.0
97.10	505.0
98.00	12249.0
99.00	10791.0
99.90	917.0
101.00	6488.0
102.00	403.0
103.00	2260.0
104.00	3985.0
105.00	3355.0
106.10	1295.0
107.00	44072.0
108.00	6933.0
109.00	1558.0
110.00	87160.0
111.00	12979.0
112.00	1777.0
113.00	625.0
115.00	478.0
116.00	2452.0
117.00	31992.0
118.00	2330.0
119.10	429.0
120.00	646.0
120.90	526.0
122.00	3183.0
123.00	4737.0
124.00	1938.0
125.00	1860.0
127.00	162688.0
128.00	12780.0
129.00	62120.0
130.00	5200.0
131.00	1236.0
132.00	694.0
133.00	451.0
133.90	1675.0
135.00	5462.0
136.00	1819.0
137.00	2713.0
138.90	543.0
140.10	539.0
141.00	7622.0

Instrument :
BNA_N
ClientSampleId :
DFTPP14

142.00	2433.0
142.90	2203.0
143.90	491.0
144.90	624.0
146.00	1461.0
146.90	4727.0
147.90	9083.0
149.00	1780.0
150.10	725.0
151.00	1332.0
151.90	923.0
152.90	3022.0
153.90	1866.0
155.00	4783.0
156.00	6342.0
156.90	1400.0
157.80	1651.0
159.00	1258.0
159.90	2538.0
161.00	3514.0
161.90	1288.0
162.90	345.0
163.90	712.0
165.00	3361.0
166.00	2582.0
167.00	15342.0
167.90	7171.0
169.00	1544.0
169.90	803.0
170.90	1060.0
171.90	1308.0
173.00	1942.0
173.90	3153.0
175.00	6567.0
176.00	2111.0
176.90	2999.0
178.00	922.0
178.90	12293.0
180.00	8404.0
181.00	3826.0
182.00	819.0
182.90	335.0
183.90	1035.0
185.00	5154.0
186.00	45440.0
187.00	13733.0
188.00	1011.0
188.90	3240.0
190.00	666.0
191.00	1687.0
192.00	4192.0
193.00	4786.0
193.90	1054.0
194.90	877.0
196.00	9957.0
197.90	355520.0
198.90	24216.0
199.90	1781.0
201.60	1348.0
202.90	2154.0
204.00	11970.0
205.00	21032.0
206.00	91032.0
207.00	12432.0
207.90	3059.0
209.00	1104.0
210.00	1665.0
210.90	3551.0
212.90	448.0
214.90	1037.0
216.00	1644.0
216.90	22680.0
218.00	3217.0
219.10	323.0
220.10	770.0
221.00	14001.0
221.80	2884.0
222.90	5539.0
224.00	48672.0
225.00	12440.0
226.00	1578.0
227.00	20200.0
228.00	3008.0
229.00	4739.0
230.00	714.0
230.90	1931.0
232.00	308.0
232.90	421.0
234.00	1439.0
235.00	1427.0
236.00	642.0
237.00	2119.0
239.00	839.0
240.00	619.0
241.00	1178.0

Instrument :
BNA_N
ClientSampleId :
DFTPP14

242.00	2484.0
242.90	2218.0
244.00	38344.0
245.00	5488.0
246.00	6936.0
247.00	1187.0
247.90	393.0
249.00	1293.0
250.90	378.0
252.00	481.0
253.00	788.0
255.00	205632.0
256.00	29776.0
257.00	2011.0
258.00	10143.0
258.90	1914.0
260.00	450.0
261.00	302.0
265.00	4224.0
266.00	834.0
266.90	405.0
270.90	494.0
271.90	664.0
273.00	5738.0
274.00	15873.0
275.00	91584.0
276.00	12134.0
277.00	7286.0
278.00	1471.0
280.90	523.0
283.00	933.0
284.10	509.0
284.90	1511.0
289.00	408.0
291.90	398.0
292.90	2035.0
293.90	341.0
296.00	22408.0
297.00	3240.0
301.00	414.0
301.90	610.0
303.00	2898.0
304.00	672.0
309.10	327.0
313.90	1019.0
314.90	2471.0
316.00	1505.0
317.20	343.0
321.00	697.0
321.90	412.0
323.00	7527.0
323.90	1808.0
326.00	316.0
326.80	1288.0
328.00	717.0
331.90	632.0
333.00	817.0
334.00	5289.0
335.00	1345.0
341.10	835.0
342.00	350.0
346.00	1346.0
346.90	321.0
352.00	2463.0
353.00	2160.0
353.90	2276.0
355.00	529.0
364.90	10528.0
366.00	1638.0
370.90	562.0
372.00	4636.0
373.00	974.0
383.00	1134.0
384.00	391.0
390.00	661.0
391.00	489.0
402.00	1653.0
403.00	2933.0
404.10	582.0
421.00	2709.0
422.00	2393.0
423.00	17104.0
424.00	3660.0
424.90	370.0
441.00	47128.0
442.00	307072.0
443.00	59944.0
444.00	5031.0

Instrument :
BNA_N
ClientSampleId :
DFTPP14

m/z	Abundance
37.00	355.0
38.10	1260.0
39.00	8424.0
40.00	918.0
41.00	990.0
42.90	601.0
44.00	6933.0
49.00	734.0
50.00	25560.0
51.00	106552.0
52.00	5725.0
53.00	416.0
55.00	1286.0
56.00	3562.0
57.00	8223.0
57.90	472.0
61.00	1231.0
62.00	1599.0
63.00	4635.0
64.10	691.0
65.00	2387.0
66.00	339.0
67.00	502.0
68.10	2378.0
69.00	112920.0
69.90	818.0
73.00	1481.0
74.00	10173.0
75.00	17544.0
76.00	5538.0
77.00	123256.0
78.00	8505.0
79.00	8030.0
80.00	6257.0
81.00	9267.0
81.90	2587.0
83.00	2320.0
83.80	392.0
85.00	1839.0
86.00	1910.0
87.00	1253.0
88.00	301.0
90.90	2335.0
92.00	2394.0
93.00	14174.0
94.10	1292.0
95.00	613.0
96.00	872.0
96.90	579.0
98.00	10060.0
98.90	8480.0
100.00	660.0
101.00	5345.0
101.90	389.0
102.90	1682.0
104.00	3291.0
105.00	3506.0
106.00	1213.0
107.00	38944.0
108.00	5632.0
109.00	1566.0
110.00	73224.0
111.00	11056.0
111.90	1213.0
113.00	625.0
114.90	326.0
116.00	2397.0
117.00	27264.0
117.90	2042.0
119.00	886.0
120.00	600.0
121.00	426.0
122.00	2667.0
123.00	4504.0
124.00	1769.0
124.90	1865.0
127.00	141568.0
128.00	10541.0
129.00	51736.0
129.90	4091.0
131.00	1096.0
132.00	598.0
132.90	458.0
133.90	1775.0
135.00	4315.0
136.00	1872.0
136.90	2187.0
137.80	642.0
139.00	408.0
139.90	674.0
141.00	8218.0

Instrument :
BNA_N
ClientSampleId :
DFTPP14

142.00	2155.0
143.00	1730.0
143.90	415.0
144.90	748.0
145.90	1156.0
146.90	4232.0
147.90	7609.0
149.00	1667.0
150.10	351.0
151.10	857.0
151.70	765.0
152.90	2294.0
153.90	1799.0
155.00	4252.0
156.00	5929.0
157.00	1260.0
157.90	1447.0
158.90	1164.0
159.90	2000.0
161.00	3239.0
162.00	986.0
162.90	492.0
164.00	532.0
165.00	2959.0
165.90	1742.0
167.00	13546.0
168.00	5333.0
169.00	1352.0
170.00	456.0
171.00	819.0
172.00	1402.0
172.90	1603.0
174.00	3253.0
175.00	5862.0
176.00	1698.0
176.90	2975.0
178.00	886.0
179.00	11649.0
179.90	7348.0
180.90	3435.0
182.00	748.0
182.90	423.0
183.90	1130.0
185.00	4856.0
186.00	42032.0
187.00	11893.0
188.00	1313.0
188.90	2745.0
189.80	776.0
191.00	1757.0
192.00	3902.0
193.00	4240.0
193.90	945.0
194.90	606.0
195.90	9436.0
196.70	1628.0
197.90	329984.0
198.90	20944.0
199.90	1811.0
201.50	1664.0
203.00	2167.0
204.00	11348.0
205.00	19120.0
206.00	86352.0
207.00	11588.0
208.00	2790.0
209.00	1058.0
210.10	1140.0
210.90	2826.0
215.00	919.0
215.90	1635.0
217.00	21704.0
217.90	2563.0
221.00	14379.0
221.90	3408.0
223.00	4597.0
224.00	47808.0
225.00	11392.0
226.00	1415.0
227.00	18896.0
228.00	2949.0
228.90	4776.0
230.00	615.0
231.00	1638.0
232.90	337.0
233.90	1229.0
234.90	978.0
235.90	965.0
236.90	1535.0
238.90	941.0
240.00	698.0
240.90	757.0
242.00	1978.0
243.00	2438.0
244.00	36496.0

Instrument :
BNA_N
ClientSampleId :
DFTPP14

245.00	4802.0
246.00	6757.0
247.00	1368.0
248.00	379.0
249.00	1582.0
252.90	936.0
255.00	195072.0
256.00	28880.0
257.00	2150.0
257.90	10737.0
258.90	1666.0
259.90	373.0
260.90	366.0
264.90	4433.0
265.90	627.0
270.80	502.0
271.90	543.0
273.00	5216.0
274.00	15348.0
275.00	94048.0
276.00	12242.0
277.00	7037.0
277.90	1271.0
281.90	435.0
283.00	1016.0
284.00	439.0
284.90	1130.0
286.00	386.0
290.10	373.0
292.10	439.0
292.90	1569.0
293.90	334.0
296.00	24584.0
297.00	2919.0
301.90	508.0
303.00	3274.0
304.00	766.0
307.90	392.0
309.00	337.0
313.80	1042.0
314.00	1015.0
314.90	2657.0
316.00	1735.0
316.90	334.0
320.90	975.0
322.10	347.0
323.00	8063.0
324.10	1616.0
327.10	1282.0
328.00	735.0
332.00	673.0
333.00	699.0
334.00	4920.0
335.00	1307.0
340.90	1032.0
342.00	391.0
345.90	1725.0
346.90	400.0
352.00	2636.0
353.00	2085.0
354.00	2833.0
355.00	491.0
364.90	11425.0
365.90	1659.0
371.00	833.0
372.00	4678.0
373.10	1132.0
383.00	1459.0
384.10	427.0
389.90	579.0
391.00	362.0
391.90	362.0
401.00	575.0
401.90	1858.0
402.90	2806.0
404.00	967.0
421.00	2593.0
422.00	2240.0
423.00	16672.0
424.00	4274.0
424.90	525.0
441.00	56424.0
442.00	358592.0
443.00	67888.0
444.00	6274.0
444.90	328.0

Instrument :
BNA_N
ClientSampleId :
DFTPP14

m/z	Abundance
39.90	720.0
41.00	566.0
43.00	938.0
44.00	6935.0
55.10	950.0
57.00	613.0
66.90	464.0
69.10	609.0
72.90	883.0
77.00	426.0
79.00	448.0
80.90	424.0
83.00	436.0
91.00	547.0
93.00	443.0
95.00	496.0
97.00	427.0
105.00	671.0
107.10	420.0
115.00	374.0
119.10	447.0
120.90	511.0
189.00	308.0
191.00	614.0
192.00	474.0
207.00	811.0
208.00	357.0

Instrument :
BNA_N
ClientSampleId :
DFTPP14