

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091019\
 Data File : BN007579.D
 Acq On : 10 Sep 2019 16:50
 Operator : HP/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN091019

Quant Time: Sep 10 17:26:05 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 10 16:53:18 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.422	0.422	0.0	140	0.00
3	n-Nitrosodimethylamine	0.239	0.304	-27.2#	99	0.00
4 S	2-Fluorophenol	1.218	1.095	10.1	101	0.00
5 S	Phenol-d6	1.870	1.432	23.4	100	0.00
6	bis(2-Chloroethyl)ether	0.789	0.652	17.4	137	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
8 S	Nitrobenzene-d5	0.244	0.249	-2.0	115	0.00
9	Naphthalene	1.125	1.101	2.1	100	0.00
10	Hexachlorobutadiene	0.226	0.222	1.8	99	0.00
11 SURR	2-Methylnaphthalene-d10	0.690	0.668	3.2	100	0.00
12	2-Methylnaphthalene	0.789	0.764	3.2	100	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.152	0.125	17.8	101	-0.01
15 S	2-Fluorobiphenyl	1.653	1.637	1.0	101	0.00
16	Acenaphthylene	2.148	2.018	6.1	98	-0.01
17	Acenaphthene	1.357	1.302	4.1	99	-0.01
18	Fluorene	1.787	1.703	4.7	98	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
20	4-Bromophenyl-phenylether	0.237	0.227	4.2	100	0.00
21	Hexachlorobenzene	0.249	0.241	3.2	99	0.00
22	Pentachlorophenol	0.069	0.056	18.8	104	0.00
23	Phenanthrene	1.235	1.186	4.0	99	0.00
24	Anthracene	1.139	1.048	8.0	97	0.00
25 SURR	Fluoranthene-d10	1.300	1.212	6.8	97	0.00
26	Fluoranthene	1.498	1.409	5.9	98	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	98	-0.01
28	Pvrene	1.438	1.398	2.8	97	0.00
29 S	Terphenyl-d14	1.041	1.011	2.9	98	0.00
30	Benzo(a)anthracene	1.431	1.310	8.5	94	0.00
31	Chrysene	1.454	1.423	2.1	100	0.00
32	Bis(2-ethylhexyl)phthalate	0.490	0.395	19.4	94	0.00
33	Indeno(1,2,3-cd)pyrene	1.538	1.460	5.1	97	0.00
34 I	Perylene-d12	1.000	1.000	0.0	97	0.00
35	Benzo(b)fluoranthene	1.422	1.367	3.9	97	0.00
36	Benzo(k)fluoranthene	1.405	1.363	3.0	99	0.00
37 C	Benzo(a)pyrene	1.265	1.194	5.6	97	0.00
38	Dibenzo(a,h)anthracene	1.270	1.217	4.2	98	0.00
39	Benzo(a,h,i)perylene	1.286	1.245	3.2	97	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091019\
Data File : BN007579.D
Acq On : 10 Sep 2019 16:50
Operator : HP/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN091019

Quant Time: Sep 10 17:26:05 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 10 16:53:18 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			