

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091119\
 Data File : BN007586.D
 Acq On : 11 Sep 2019 13:47
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 11 16:20:31 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 11 14:14:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.422	0.424	-0.5	147	0.00
3	n-Nitrosodimethylamine	0.239	0.330	-38.1#	112	0.00
4 S	2-Fluorophenol	1.218	1.306	-7.2	126	0.00
5 S	Phenol-d6	1.870	1.677	10.3	123	0.00
6	bis(2-Chloroethyl)ether	0.789	1.116	-41.4#	245#	-0.03
7 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
8 S	Nitrobenzene-d5	0.244	0.359	-47.1#	168#	0.00
9	Naphthalene	1.125	1.204	-7.0	111	0.00
10	Hexachlorobutadiene	0.226	0.254	-12.4	114	0.00
11 SURR	2-Methylnaphthalene-d10	0.690	0.732	-6.1	111	0.00
12	2-Methylnaphthalene	0.789	0.834	-5.7	110	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	106	0.00
14 S	2,4,6-Tribromophenol	0.152	0.176	-15.8	152#	-0.01
15 S	2-Fluorobiphenyl	1.653	1.659	-0.4	110	-0.01
16	Acenaphthylene	2.148	2.102	2.1	110	-0.01
17	Acenaphthene	1.357	1.320	2.7	108	-0.01
18	Fluorene	1.787	1.739	2.7	108	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
20	4-Bromophenyl-phenylether	0.237	0.261	-10.1	110	0.00
21	Hexachlorobenzene	0.249	0.278	-11.6	110	0.00
22	Pentachlorophenol	0.069	0.079	-14.5	141	0.00
23	Phenanthrene	1.235	1.323	-7.1	106	-0.01
24	Anthracene	1.139	1.232	-8.2	109	0.00
25 SURR	Fluoranthene-d10	1.300	1.439	-10.7	111	0.00
26	Fluoranthene	1.498	1.642	-9.6	110	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	104	-0.01
28	Pvrene	1.438	1.510	-5.0	111	0.00
29 S	Terphenyl-d14	1.041	1.113	-6.9	113	0.00
30	Benzo(a)anthracene	1.431	1.557	-8.8	117	0.00
31	Chrysene	1.454	1.553	-6.8	114	-0.01
32	Bis(2-ethylhexyl)phthalate	0.490	0.726	-48.2#	182#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.538	1.844	-19.9	128	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	119	0.00
35	Benzo(b)fluoranthene	1.422	1.376	3.2	120	0.00
36	Benzo(k)fluoranthene	1.405	1.465	-4.3	131	0.00
37 C	Benzo(a)pyrene	1.265	1.259	0.5	126	-0.01
38	Dibenzo(a,h)anthracene	1.270	1.321	-4.0	130	-0.01
39	Benzo(a,h,i)perylene	1.286	1.282	0.3	123	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091119\
Data File : BN007586.D
Acq On : 11 Sep 2019 13:47
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 11 16:20:31 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 11 14:14:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			