

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091123\
 Data File : BN027567.D
 Acq On : 11 Sep 2023 08:36
 Operator : MA/JU
 Sample : 04251-04
 Misc :
 ALS Vial : 78 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 OW-07B-65-083023-3V

Quant Time: Sep 12 01:29:44 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 11 05:26:09 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/12/2023
 Supervised By :mohammad ahmed 09/13/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.037	152	6251	0.400	ng	0.00
7) Naphthalene-d8	10.853	136	20002	0.400	ng	0.00
13) Acenaphthene-d10	14.662	164	11556	0.400	ng	0.00
19) Phenanthrene-d10	17.418	188	25417m	0.400	ng	0.00
29) Chrysene-d12	21.596	240	16161	0.400	ng	# 0.00
35) Perylene-d12	24.098	264	13359	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.560	112	4538	0.278	ng	0.00
5) Phenol-d6	7.178	99	3317	0.167	ng	0.00
8) Nitrobenzene-d5	9.219	82	6700	0.460	ng	0.00
11) 2-Methylnaphthalene-d10	12.430	152	12466	0.424	ng	0.00
14) 2,4,6-Tribromophenol	16.152	330	2520	0.589	ng	0.00
15) 2-Fluorobiphenyl	13.293	172	21492	0.489	ng	0.00
27) Fluoranthene-d10	19.435	212	27374m	0.433	ng	0.00
31) Terphenyl-d14	20.020	244	24268	0.748	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.436	88	172231	22.570	ng	Qvalue 98
34) Bis(2-ethylhexyl)phtha...	21.444	149	1223	0.046	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091123\
Data File : BN027567.D
Acq On : 11 Sep 2023 08:36
Operator : MA/JU
Sample : 04251-04
Misc :
ALS Vial : 78 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
OW-07B-65-083023-3V

Quant Time: Sep 12 01:29:44 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 11 05:26:09 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/12/2023
Supervised By :mohammad ahmed 09/13/2023

