

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091123\
 Data File : BN027572.D
 Acq On : 11 Sep 2023 11:41
 Operator : MA/JU
 Sample : 04252-05
 Misc :
 ALS Vial : 83 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW-03-6-083023

Quant Time: Sep 12 01:31:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 11 05:26:09 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.037	152	6035	0.400	ng	0.00
7) Naphthalene-d8	10.853	136	18303	0.400	ng	0.00
13) Acenaphthene-d10	14.662	164	11355	0.400	ng	0.00
19) Phenanthrene-d10	17.405	188	24023	0.400	ng	-0.01
29) Chrysene-d12	21.587	240	15437	0.400	ng	0.00
35) Perylene-d12	24.095	264	13028	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.560	112	3562	0.226	ng	0.00
5) Phenol-d6	7.185	99	2590	0.135	ng	0.00
8) Nitrobenzene-d5	9.230	82	5778	0.433	ng	0.01
11) 2-Methylnaphthalene-d10	12.430	152	11240	0.418	ng	0.00
14) 2,4,6-Tribromophenol	16.152	330	2418	0.575	ng	0.00
15) 2-Fluorobiphenyl	13.293	172	20074	0.465	ng	0.00
27) Fluoranthene-d10	19.435	212	26225	0.439	ng	0.00
31) Terphenyl-d14	20.020	244	22673	0.731	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.444	149	3476	0.136	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091123\
Data File : BN027572.D
Acq On : 11 Sep 2023 11:41
Operator : MA/JU
Sample : 04252-05
Misc :
ALS Vial : 83 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW-03-6-083023

Quant Time: Sep 12 01:31:05 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 11 05:26:09 2023
Response via : Initial Calibration

