




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091023\
Data File : BN027560.D
Acq On : 11 Sep 2023 04:19
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 71 Sample Multiplier: 1

Quant Time: Sep 11 05:05:07 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 11 05:05:02 2023
Response via : Initial Calibration



TIC: BN027560.D\data.ms

(70) Pentachlorophenol (C)




17.039min (0.000) 18876.54 ng

response 483968

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.35
264.00	61.60	58.95
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091023\
Data File : BN027560.D
Acq On : 11 Sep 2023 04:19
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 71 Sample Multiplier: 1

Quant Time: Sep 11 05:05:07 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 11 05:05:02 2023
Response via : Initial Calibration



TIC: BN027560.D\data.ms

(77) Benzidine

19.675min (0.000) 0.00 ng

response 3000358

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.30
183.00	13.20	11.56
0.00	0.00	0.00