

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091219\
 Data File : BN007608.D
 Acq On : 12 Sep 2019 14:40
 Operator : HP/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN091219

Quant Time: Sep 12 15:39:26 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 12 14:27:51 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	62	0.00
2	1,4-Dioxane	0.479	0.475	0.8	64	0.00
3	n-Nitrosodimethylamine	0.217	0.328	-51.2#	66	0.00
4 S	2-Fluorophenol	1.281	1.137	11.2	60	0.00
5 S	Phenol-d6	1.803	1.446	19.8	59	0.00
6	bis(2-Chloroethyl)ether	1.174	1.092	7.0	57	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	61	0.00
8 S	Nitrobenzene-d5	0.319	0.319	0.0	59	0.00
9	Naphthalene	1.146	1.219	-6.4	62	-0.01
10	Hexachlorobutadiene	0.231	0.246	-6.5	62	-0.01
11 SURR	2-Methylnaphthalene-d10	0.677	0.709	-4.7	61	0.00
12	2-Methylnaphthalene	0.770	0.813	-5.6	61	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	61	0.00
14 S	2,4,6-Tribromophenol	0.151	0.123	18.5	58	0.00
15 S	2-Fluorobiphenyl	1.745	1.754	-0.5	62	0.00
16	Acenaphthylene	2.081	1.959	5.9	59	0.00
17	Acenaphthene	1.357	1.341	1.2	61	0.00
18	Fluorene	1.712	1.674	2.2	61	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	61	0.00
20	4-Bromophenyl-phenylether	0.247	0.261	-5.7	61	0.00
21	Hexachlorobenzene	0.264	0.287	-8.7	62	0.00
22	Pentachlorophenol	0.069	0.058	15.9	56	0.00
23	Phenanthrene	1.262	1.348	-6.8	61	0.00
24	Anthracene	1.119	1.151	-2.9	60	0.00
25 SURR	Fluoranthene-d10	1.262	1.322	-4.8	61	0.00
26	Fluoranthene	1.474	1.546	-4.9	60	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	61	0.00
28	Pvrene	1.453	1.534	-5.6	61	0.00
29 S	Terphenyl-d14	1.055	1.111	-5.3	62	0.00
30	Benzo(a)anthracene	1.419	1.420	-0.1	59	0.00
31	Chrysene	1.476	1.542	-4.5	61	0.00
32	Bis(2-ethylhexyl)phthalate	0.481	0.412	14.3	55	0.00
33	Indeno(1,2,3-cd)pyrene	1.576	1.648	-4.6	61	0.00
34 I	Perylene-d12	1.000	1.000	0.0	60	0.00
35	Benzo(b)fluoranthene	1.453	1.452	0.1	59	0.00
36	Benzo(k)fluoranthene	1.487	1.524	-2.5	62	0.00
37 C	Benzo(a)pyrene	1.289	1.266	1.8	60	0.00
38	Dibenzo(a,h)anthracene	1.350	1.380	-2.2	61	0.00
39	Benzo(a,h,i)perylene	1.336	1.352	-1.2	61	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091219\
Data File : BN007608.D
Acq On : 12 Sep 2019 14:40
Operator : HP/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN091219

Quant Time: Sep 12 15:39:26 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 12 14:27:51 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			