

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091219\  
 Data File : BN007608.D  
 Acq On : 12 Sep 2019 14:40  
 Operator : HP/JU  
 Sample : SSTDICV0.4  
 Misc :  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 ICVBN091219

Quant Time: Sep 12 15:39:26 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091219.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Sep 12 14:27:51 2019  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|         | Compound                   | Amount | Calc. | %Dev   | Area% | Dev(min) |
|---------|----------------------------|--------|-------|--------|-------|----------|
| 1 I     | 1,4-Dichlorobenzene-d4     | 0.400  | 0.400 | 0.0    | 62    | 0.00     |
| 2       | 1,4-Dioxane                | 0.400  | 0.397 | 0.8    | 64    | 0.00     |
| 3       | n-Nitrosodimethylamine     | 0.400  | 0.604 | -51.0# | 66    | 0.00     |
| 4 S     | 2-Fluorophenol             | 0.400  | 0.355 | 11.3   | 60    | 0.00     |
| 5 S     | Phenol-d6                  | 0.400  | 0.371 | 7.3    | 59    | 0.00     |
| 6       | bis(2-Chloroethyl)ether    | 0.400  | 0.372 | 7.0    | 57    | 0.00     |
| 7 I     | Naphthalene-d8             | 0.400  | 0.400 | 0.0    | 61    | 0.00     |
| 8 S     | Nitrobenzene-d5            | 0.400  | 0.400 | 0.0    | 59    | 0.00     |
| 9       | Naphthalene                | 0.400  | 0.426 | -6.5   | 62    | -0.01    |
| 10      | Hexachlorobutadiene        | 0.400  | 0.427 | -6.7   | 62    | -0.01    |
| 11 SURR | 2-Methylnaphthalene-d10    | 0.400  | 0.419 | -4.7   | 61    | 0.00     |
| 12      | 2-Methylnaphthalene        | 0.400  | 0.422 | -5.5   | 61    | 0.00     |
| 13 I    | Acenaphthene-d10           | 0.400  | 0.400 | 0.0    | 61    | 0.00     |
| 14 S    | 2,4,6-Tribromophenol       | 0.400  | 0.325 | 18.8   | 58    | 0.00     |
| 15 S    | 2-Fluorobiphenyl           | 0.400  | 0.402 | -0.5   | 62    | 0.00     |
| 16      | Acenaphthylene             | 0.400  | 0.377 | 5.8    | 59    | 0.00     |
| 17      | Acenaphthene               | 0.400  | 0.395 | 1.3    | 61    | 0.00     |
| 18      | Fluorene                   | 0.400  | 0.391 | 2.3    | 61    | 0.00     |
| 19 I    | Phenanthrene-d10           | 0.400  | 0.400 | 0.0    | 61    | 0.00     |
| 20      | 4-Bromophenyl-phenylether  | 0.400  | 0.422 | -5.5   | 61    | 0.00     |
| 21      | Hexachlorobenzene          | 0.400  | 0.434 | -8.5   | 62    | 0.00     |
| 22      | Pentachlorophenol          | 0.400  | 0.338 | 15.5   | 56    | 0.00     |
| 23      | Phenanthrene               | 0.400  | 0.427 | -6.7   | 61    | 0.00     |
| 24      | Anthracene                 | 0.400  | 0.412 | -3.0   | 60    | 0.00     |
| 25 SURR | Fluoranthene-d10           | 0.400  | 0.419 | -4.7   | 61    | 0.00     |
| 26      | Fluoranthene               | 0.400  | 0.419 | -4.7   | 60    | 0.00     |
| 27 I    | Chrysene-d12               | 0.400  | 0.400 | 0.0    | 61    | 0.00     |
| 28      | Pvrene                     | 0.400  | 0.422 | -5.5   | 61    | 0.00     |
| 29 S    | Terphenyl-d14              | 0.400  | 0.421 | -5.2   | 62    | 0.00     |
| 30      | Benzo(a)anthracene         | 0.400  | 0.400 | 0.0    | 59    | 0.00     |
| 31      | Chrysene                   | 0.400  | 0.418 | -4.5   | 61    | 0.00     |
| 32      | Bis(2-ethylhexyl)phthalate | 0.400  | 0.343 | 14.2   | 55    | 0.00     |
| 33      | Indeno(1,2,3-cd)pyrene     | 0.400  | 0.418 | -4.5   | 61    | 0.00     |
| 34 I    | Perylene-d12               | 0.400  | 0.400 | 0.0    | 60    | 0.00     |
| 35      | Benzo(b)fluoranthene       | 0.400  | 0.400 | 0.0    | 59    | 0.00     |
| 36      | Benzo(k)fluoranthene       | 0.400  | 0.410 | -2.5   | 62    | 0.00     |
| 37 C    | Benzo(a)pyrene             | 0.400  | 0.393 | 1.8    | 60    | 0.00     |
| 38      | Dibenzo(a,h)anthracene     | 0.400  | 0.409 | -2.2   | 61    | 0.00     |
| 39      | Benzo(a,h,i)perylene       | 0.400  | 0.405 | -1.3   | 61    | 0.00     |

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091219\  
Data File : BN007608.D  
Acq On : 12 Sep 2019 14:40  
Operator : HP/JU  
Sample : SSTDICV0.4  
Misc :  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
ICVBN091219

Quant Time: Sep 12 15:39:26 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091219.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Sep 12 14:27:51 2019  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound           | Amount                       | Calc. | %Dev Area | % Dev(min) |
|--------------------|------------------------------|-------|-----------|------------|
| -----              |                              |       |           |            |
| (#) = Out of Range | SPCC's out = 0 CCC's out = 0 |       |           |            |