

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091223\
 Data File : BN027584.D
 Acq On : 11 Sep 2023 20:21
 Operator : MA/JU
 Sample : PB155158BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK158

Quant Time: Sep 12 02:14:06 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 12 02:09:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.038	152	6387	0.400	ng/ul	0.00
4) Naphthalene-d8	10.856	136	17428	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.666	164	8146	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.413	188	15489	0.400	ng/ul	0.00
17) Chrysene-d12	21.592	240	10860	0.400	ng/ul	0.00
23) Perylene-d12	24.091	264	11717	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.439	96	49045	5.773	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.434	152	8607	0.325	ng/ul	0.00
18) Fluoranthene-d10	19.436	212	13840	0.372	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091223\
Data File : BN027584.D
Acq On : 11 Sep 2023 20:21
Operator : MA/JU
Sample : PB155158BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK158

Quant Time: Sep 12 02:14:06 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 12 02:09:21 2023
Response via : Initial Calibration

