

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091223\
 Data File : BN027588.D
 Acq On : 11 Sep 2023 22:50
 Operator : MA/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4262

Quant Time: Sep 12 02:14:47 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 12 02:09:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.038	152	6148	0.400	ng/ul	0.00
4) Naphthalene-d8	10.856	136	17030	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.666	164	7662	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.413	188	15017	0.400	ng/ul	0.00
17) Chrysene-d12	21.592	240	11321	0.400	ng/ul	0.00
23) Perylene-d12	24.094	264	10959	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.439	96	3320	0.406	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.434	152	9989	0.386	ng/ul	0.00
18) Fluoranthene-d10	19.436	212	15801	0.407	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.477	88	3283	0.397	ng/ul	99
5) Naphthalene	10.905	128	20233	0.410	ng/ul	100
7) 2-Methylnaphthalene	12.506	142	11666	0.378	ng/ul	100
8) 1-Methylnaphthalene	12.726	142	11935	0.362	ng/ul	100
10) Acenaphthylene	14.393	152	14219	0.405	ng/ul	100
11) Acenaphthene	14.731	153	11944	0.396	ng/ul	97
12) Fluorene	15.716	166	12648	0.374	ng/ul	99
14) Pentachlorophenol	17.046	266	1232	0.421	ng/ul	99
15) Phenanthrene	17.456	178	19364	0.396	ng/ul	100
16) Anthracene	17.549	178	16143	0.415	ng/ul	100
19) Fluoranthene	19.464	202	20365	0.391	ng/ul	99
20) Pyrene	19.831	202	20891	0.394	ng/ul	98
21) Benzo(a)anthracene	21.574	228	17811	0.437	ng/ul	99
22) Chrysene	21.630	228	18646	0.392	ng/ul	99
24) Benzo(b)fluoranthene	23.316	252	18833	0.364	ng/ul	97
25) Benzo(k)fluoranthene	23.369	252	18619	0.368	ng/ul	97
26) Benzo(a)pyrene	23.983	252	15746	0.390	ng/ul	97
27) Indeno(1,2,3-cd)pyrene	26.733	276	20810	0.463	ng/ul#	97
28) Dibenzo(a,h)anthracene	26.757	278	16651	0.485	ng/ul	97
29) Benzo(g,h,i)perylene	27.552	276	18168	0.448	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091223\
Data File : BN027588.D
Acq On : 11 Sep 2023 22:50
Operator : MA/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4262

Quant Time: Sep 12 02:14:47 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 12 02:09:21 2023
Response via : Initial Calibration

