

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091319\
 Data File : BN007625.D
 Acq On : 13 Sep 2019 14:03
 Operator : HP/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDC00.4

Quant Time: Sep 14 02:19:18 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 12 14:27:51 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	69	0.00
2	1,4-Dioxane	0.479	0.467	2.5	70	0.00
3	n-Nitrosodimethylamine	0.217	0.320	-47.5#	71	0.00
4 S	2-Fluorophenol	1.281	1.111	13.3	65	0.00
5 S	Phenol-d6	1.803	1.434	20.5	65	0.00
6	bis(2-Chloroethyl)ether	1.174	1.112	5.3	64	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	68	-0.01
8 S	Nitrobenzene-d5	0.319	0.309	3.1	64	0.00
9	Naphthalene	1.146	1.225	-6.9	69	-0.01
10	Hexachlorobutadiene	0.231	0.240	-3.9	67	-0.01
11 SURR	2-Methylnaphthalene-d10	0.677	0.706	-4.3	68	0.00
12	2-Methylnaphthalene	0.770	0.809	-5.1	68	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	68	0.00
14 S	2,4,6-Tribromophenol	0.151	0.115	23.8	61	0.00
15 S	2-Fluorobiphenyl	1.745	1.747	-0.1	68	0.00
16	Acenaphthylene	2.081	1.911	8.2	64	0.00
17	Acenaphthene	1.357	1.301	4.1	66	0.00
18	Fluorene	1.712	1.633	4.6	66	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	67	0.00
20	4-Bromophenyl-phenylether	0.247	0.261	-5.7	67	0.00
21	Hexachlorobenzene	0.264	0.284	-7.6	67	-0.01
22	Pentachlorophenol	0.069	0.057	17.4	61	-0.01
23	Phenanthrene	1.262	1.344	-6.5	67	0.00
24	Anthracene	1.119	1.124	-0.4	64	0.00
25 SURR	Fluoranthene-d10	1.262	1.292	-2.4	65	0.00
26	Fluoranthene	1.474	1.521	-3.2	65	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	65	-0.01
28	Pvrene	1.453	1.547	-6.5	65	0.00
29 S	Terphenyl-d14	1.055	1.115	-5.7	65	0.00
30	Benzo(a)anthracene	1.419	1.368	3.6	60	0.00
31	Chrysene	1.476	1.592	-7.9	67	0.00
32	Bis(2-ethylhexyl)phthalate	0.481	0.395	17.9	56	0.00
33	Indeno(1,2,3-cd)pyrene	1.576	1.705	-8.2	66	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	64	0.00
35	Benzo(b)fluoranthene	1.453	1.467	-1.0	64	0.00
36	Benzo(k)fluoranthene	1.487	1.535	-3.2	67	0.00
37 C	Benzo(a)pyrene	1.289	1.246	3.3	63	0.00
38	Dibenzo(a,h)anthracene	1.350	1.382	-2.4	65	0.00
39	Benzo(a,h,i)perylene	1.336	1.341	-0.4	65	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091319\
Data File : BN007625.D
Acq On : 13 Sep 2019 14:03
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 14 02:19:18 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 12 14:27:51 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			