

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091319\
 Data File : BN007625.D
 Acq On : 13 Sep 2019 14:03
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 14 02:19:18 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 12 14:27:51 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	69	0.00
2	1,4-Dioxane	0.400	0.390	2.5	70	0.00
3	n-Nitrosodimethylamine	0.400	0.589	-47.2#	71	0.00
4 S	2-Fluorophenol	0.400	0.347	13.3	65	0.00
5 S	Phenol-d6	0.400	0.368	8.0	65	0.00
6	bis(2-Chloroethyl)ether	0.400	0.379	5.3	64	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	68	-0.01
8 S	Nitrobenzene-d5	0.400	0.387	3.3	64	0.00
9	Naphthalene	0.400	0.427	-6.7	69	-0.01
10	Hexachlorobutadiene	0.400	0.416	-4.0	67	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.417	-4.2	68	0.00
12	2-Methylnaphthalene	0.400	0.420	-5.0	68	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	68	0.00
14 S	2,4,6-Tribromophenol	0.400	0.305	23.8	61	0.00
15 S	2-Fluorobiphenyl	0.400	0.401	-0.3	68	0.00
16	Acenaphthylene	0.400	0.367	8.3	64	0.00
17	Acenaphthene	0.400	0.384	4.0	66	0.00
18	Fluorene	0.400	0.382	4.5	66	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	67	0.00
20	4-Bromophenyl-phenylether	0.400	0.423	-5.7	67	0.00
21	Hexachlorobenzene	0.400	0.430	-7.5	67	-0.01
22	Pentachlorophenol	0.400	0.333	16.8	61	-0.01
23	Phenanthrene	0.400	0.426	-6.5	67	0.00
24	Anthracene	0.400	0.402	-0.5	64	0.00
25 SURR	Fluoranthene-d10	0.400	0.409	-2.2	65	0.00
26	Fluoranthene	0.400	0.413	-3.2	65	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	65	-0.01
28	Pvrene	0.400	0.426	-6.5	65	0.00
29 S	Terphenyl-d14	0.400	0.423	-5.7	65	0.00
30	Benzo(a)anthracene	0.400	0.385	3.8	60	0.00
31	Chrysene	0.400	0.432	-8.0	67	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.329	17.8	56	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.433	-8.2	66	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	64	0.00
35	Benzo(b)fluoranthene	0.400	0.404	-1.0	64	0.00
36	Benzo(k)fluoranthene	0.400	0.413	-3.2	67	0.00
37 C	Benzo(a)pyrene	0.400	0.387	3.3	63	0.00
38	Dibenzo(a,h)anthracene	0.400	0.409	-2.2	65	0.00
39	Benzo(a,h,i)perylene	0.400	0.402	-0.5	65	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091319\
Data File : BN007625.D
Acq On : 13 Sep 2019 14:03
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 14 02:19:18 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 12 14:27:51 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				