

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091319\
 Data File : BN007627.D
 Acq On : 13 Sep 2019 15:15
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 14 02:59:54 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 12 14:27:51 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	67	0.00
2	1,4-Dioxane	0.479	0.461	3.8	68	0.00
3	n-Nitrosodimethylamine	0.217	0.316	-45.6#	69	0.00
4 S	2-Fluorophenol	1.281	1.072	16.3	62	0.00
5 S	Phenol-d6	1.803	1.387	23.1	62	0.00
6	bis(2-Chloroethyl)ether	1.174	1.064	9.4	61	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	67	-0.01
8 S	Nitrobenzene-d5	0.319	0.306	4.1	62	0.00
9	Naphthalene	1.146	1.220	-6.5	67	-0.01
10	Hexachlorobutadiene	0.231	0.242	-4.8	66	-0.01
11 SURR	2-Methylnaphthalene-d10	0.677	0.703	-3.8	66	0.00
12	2-Methylnaphthalene	0.770	0.807	-4.8	66	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	66	0.00
14 S	2,4,6-Tribromophenol	0.151	0.112	25.8#	57	0.00
15 S	2-Fluorobiphenyl	1.745	1.763	-1.0	67	0.00
16	Acenaphthylene	2.081	1.887	9.3	61	0.00
17	Acenaphthene	1.357	1.317	2.9	65	0.00
18	Fluorene	1.712	1.661	3.0	65	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	64	0.00
20	4-Bromophenyl-phenylether	0.247	0.265	-7.3	65	0.00
21	Hexachlorobenzene	0.264	0.286	-8.3	65	-0.01
22	Pentachlorophenol	0.069	0.059	14.5	60	-0.01
23	Phenanthrene	1.262	1.350	-7.0	64	0.00
24	Anthracene	1.119	1.115	0.4	61	0.00
25 SURR	Fluoranthene-d10	1.262	1.286	-1.9	62	0.00
26	Fluoranthene	1.474	1.520	-3.1	62	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	62	-0.01
28	Pvrene	1.453	1.553	-6.9	63	0.00
29 S	Terphenyl-d14	1.055	1.115	-5.7	63	0.00
30	Benzo(a)anthracene	1.419	1.348	5.0	57	0.00
31	Chrysene	1.476	1.592	-7.9	64	0.00
32	Bis(2-ethylhexyl)phthalate	0.481	0.376	21.8	51	-0.01
33	Indeno(1,2,3-cd)pyrene	1.576	1.735	-10.1	65	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	62	0.00
35	Benzo(b)fluoranthene	1.453	1.449	0.3	61	0.00
36	Benzo(k)fluoranthene	1.487	1.543	-3.8	65	0.00
37 C	Benzo(a)pyrene	1.289	1.224	5.0	60	-0.01
38	Dibenzo(a,h)anthracene	1.350	1.404	-4.0	64	-0.01
39	Benzo(a,h,i)perylene	1.336	1.341	-0.4	62	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091319\
Data File : BN007627.D
Acq On : 13 Sep 2019 15:15
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 14 02:59:54 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 12 14:27:51 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			