

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091319\
 Data File : BN007643.D
 Acq On : 14 Sep 2019 00:48
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 14 05:20:49 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 12 14:27:51 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	72	0.00
2	1,4-Dioxane	0.479	0.458	4.4	72	0.00
3	n-Nitrosodimethylamine	0.217	0.332	-53.0#	78	0.00
4 S	2-Fluorophenol	1.281	1.048	18.2	65	0.00
5 S	Phenol-d6	1.803	1.339	25.7#	64	0.00
6	bis(2-Chloroethyl)ether	1.174	1.094	6.8	67	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	72	-0.01
8 S	Nitrobenzene-d5	0.319	0.302	5.3	65	-0.01
9	Naphthalene	1.146	1.225	-6.9	72	-0.01
10	Hexachlorobutadiene	0.231	0.239	-3.5	70	-0.01
11 SURR	2-Methylnaphthalene-d10	0.677	0.706	-4.3	71	0.00
12	2-Methylnaphthalene	0.770	0.810	-5.2	71	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	71	0.00
14 S	2,4,6-Tribromophenol	0.151	0.114	24.5	62	0.00
15 S	2-Fluorobiphenyl	1.745	1.792	-2.7	73	0.00
16	Acenaphthylene	2.081	1.861	10.6	65	0.00
17	Acenaphthene	1.357	1.333	1.8	70	0.00
18	Fluorene	1.712	1.664	2.8	70	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	70	0.00
20	4-Bromophenyl-phenylether	0.247	0.262	-6.1	70	0.00
21	Hexachlorobenzene	0.264	0.282	-6.8	70	0.00
22	Pentachlorophenol	0.069	0.055	20.3	62	0.00
23	Phenanthrene	1.262	1.345	-6.6	70	0.00
24	Anthracene	1.119	1.104	1.3	66	0.00
25 SURR	Fluoranthene-d10	1.262	1.268	-0.5	67	0.00
26	Fluoranthene	1.474	1.504	-2.0	67	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	67	-0.02
28	Pvrene	1.453	1.551	-6.7	67	0.00
29 S	Terphenyl-d14	1.055	1.129	-7.0	68	0.00
30	Benzo(a)anthracene	1.419	1.347	5.1	61	0.00
31	Chrysene	1.476	1.613	-9.3	70	0.00
32	Bis(2-ethylhexyl)phthalate	0.481	0.372	22.7	55	0.00
33	Indeno(1,2,3-cd)pyrene	1.576	1.852	-17.5	74	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	68	-0.02
35	Benzo(b)fluoranthene	1.453	1.455	-0.1	68	-0.02
36	Benzo(k)fluoranthene	1.487	1.575	-5.9	72	-0.02
37 C	Benzo(a)pyrene	1.289	1.241	3.7	66	-0.02
38	Dibenzo(a,h)anthracene	1.350	1.451	-7.5	73	-0.02
39	Benzo(a,h,i)perylene	1.336	1.371	-2.6	70	-0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091319\
Data File : BN007643.D
Acq On : 14 Sep 2019 00:48
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 14 05:20:49 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 12 14:27:51 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			