

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091319\
 Data File : BN007645.D
 Acq On : 14 Sep 2019 03:11
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 16 01:36:38 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 12 14:27:51 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
2	1,4-Dioxane	0.479	0.418	12.7	72	0.00
3	n-Nitrosodimethylamine	0.217	0.335	-54.4#	86	0.00
4 S	2-Fluorophenol	1.281	1.067	16.7	72	0.00
5 S	Phenol-d6	1.803	1.384	23.2	73	0.00
6	bis(2-Chloroethyl)ether	1.174	1.102	6.1	74	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	78	-0.01
8 S	Nitrobenzene-d5	0.319	0.305	4.4	72	-0.01
9	Naphthalene	1.146	1.222	-6.6	79	-0.01
10	Hexachlorobutadiene	0.231	0.238	-3.0	76	-0.01
11 SURR	2-Methylnaphthalene-d10	0.677	0.702	-3.7	77	0.00
12	2-Methylnaphthalene	0.770	0.810	-5.2	78	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	77	0.00
14 S	2,4,6-Tribromophenol	0.151	0.113	25.2#	68	0.00
15 S	2-Fluorobiphenyl	1.745	1.786	-2.3	79	0.00
16	Acenaphthylene	2.081	1.857	10.8	70	0.00
17	Acenaphthene	1.357	1.332	1.8	77	0.00
18	Fluorene	1.712	1.657	3.2	76	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	76	0.00
20	4-Bromophenyl-phenylether	0.247	0.263	-6.5	76	0.00
21	Hexachlorobenzene	0.264	0.284	-7.6	76	0.00
22	Pentachlorophenol	0.069	0.061	11.6	73	-0.01
23	Phenanthrene	1.262	1.350	-7.0	76	0.00
24	Anthracene	1.119	1.111	0.7	72	0.00
25 SURR	Fluoranthene-d10	1.262	1.265	-0.2	73	0.00
26	Fluoranthene	1.474	1.516	-2.8	74	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	73	-0.02
28	Pvrene	1.453	1.554	-7.0	74	0.00
29 S	Terphenyl-d14	1.055	1.133	-7.4	75	0.00
30	Benzo(a)anthracene	1.419	1.357	4.4	67	0.00
31	Chrysene	1.476	1.631	-10.5	77	0.00
32	Bis(2-ethylhexyl)phthalate	0.481	0.373	22.5	60	-0.02
33	Indeno(1,2,3-cd)pyrene	1.576	1.853	-17.6	81	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	74	-0.02
35	Benzo(b)fluoranthene	1.453	1.418	2.4	72	-0.02
36	Benzo(k)fluoranthene	1.487	1.644	-10.6	83	-0.02
37 C	Benzo(a)pyrene	1.289	1.226	4.9	72	-0.02
38	Dibenzo(a,h)anthracene	1.350	1.450	-7.4	79	-0.03
39	Benzo(a,h,i)perylene	1.336	1.370	-2.5	77	-0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091319\
Data File : BN007645.D
Acq On : 14 Sep 2019 03:11
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 16 01:36:38 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 12 14:27:51 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			