

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091319\
 Data File : BN007660.D
 Acq On : 14 Sep 2019 12:09
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 16 01:39:04 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 12 14:27:51 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
2	1,4-Dioxane	0.479	0.454	5.2	79	0.00
3	n-Nitrosodimethylamine	0.217	0.304	-40.1#	79	0.00
4 S	2-Fluorophenol	1.281	1.133	11.6	77	0.00
5 S	Phenol-d6	1.803	1.471	18.4	78	0.00
6	bis(2-Chloroethyl)ether	1.174	1.091	7.1	74	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	80	-0.01
8 S	Nitrobenzene-d5	0.319	0.321	-0.6	77	-0.01
9	Naphthalene	1.146	1.219	-6.4	80	-0.01
10	Hexachlorobutadiene	0.231	0.240	-3.9	78	-0.01
11 SURR	2-Methylnaphthalene-d10	0.677	0.702	-3.7	79	0.00
12	2-Methylnaphthalene	0.770	0.812	-5.5	80	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	80	0.00
14 S	2,4,6-Tribromophenol	0.151	0.120	20.5	74	-0.01
15 S	2-Fluorobiphenyl	1.745	1.761	-0.9	81	-0.01
16	Acenaphthylene	2.081	1.903	8.6	75	-0.01
17	Acenaphthene	1.357	1.333	1.8	79	-0.01
18	Fluorene	1.712	1.655	3.3	78	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	78	-0.01
20	4-Bromophenyl-phenylether	0.247	0.261	-5.7	78	-0.01
21	Hexachlorobenzene	0.264	0.280	-6.1	77	-0.01
22	Pentachlorophenol	0.069	0.061	11.6	76	-0.01
23	Phenanthrene	1.262	1.365	-8.2	79	-0.01
24	Anthracene	1.119	1.124	-0.4	76	0.00
25 SURR	Fluoranthene-d10	1.262	1.284	-1.7	76	-0.01
26	Fluoranthene	1.474	1.530	-3.8	77	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	77	-0.01
28	Pvrene	1.453	1.537	-5.8	77	-0.01
29 S	Terphenyl-d14	1.055	1.105	-4.7	77	-0.01
30	Benzo(a)anthracene	1.419	1.404	1.1	73	-0.01
31	Chrysene	1.476	1.570	-6.4	79	-0.01
32	Bis(2-ethylhexyl)phthalate	0.481	0.409	15.0	70	-0.01
33	Indeno(1,2,3-cd)pyrene	1.576	1.805	-14.5	84	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	77	-0.01
35	Benzo(b)fluoranthene	1.453	1.481	-1.9	78	-0.01
36	Benzo(k)fluoranthene	1.487	1.603	-7.8	84	-0.01
37 C	Benzo(a)pyrene	1.289	1.301	-0.9	79	-0.01
38	Dibenzo(a,h)anthracene	1.350	1.429	-5.9	82	-0.02
39	Benzo(a,h,i)perylene	1.336	1.369	-2.5	80	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091319\
Data File : BN007660.D
Acq On : 14 Sep 2019 12:09
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 16 01:39:04 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 12 14:27:51 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			